

# **MEDICAL LABORATORY**

# **EVALUATION**

## **PARTICIPANT SUMMARY**

# **2 • 0 • 1 • 8**

**Please see the corresponding US participant summary for any statistics not represented in this supplement.**

**International Data Supplement  
2018 MLE-M2**

**ACP | Medical Laboratory  
Evaluation** 

Total Commitment to Education and Service  
Provided by ACP, Inc.

**COPYRIGHT © ACP, Inc.**

*Single copies of all material subject to this copyright may be photocopied for the non-commercial use of scientific or educational advancement.*

# Table of Contents

<b>Evaluation Criteria</b> .....	<b>4</b>
<b>Hematology</b>	
<b>Hematology with 5-part Automated Differential (CL Samples – Module 223)</b> .....	<b>5</b>
White Blood Cell Count .....	5
Red Blood Cell Count .....	6
Hemoglobin .....	7
Hematocrit .....	8
Platelet Count .....	9
Automated Differential .....	10
<b>Blood Bank</b>	
<b>ABO Group</b> .....	15
<b>Rh Factor (D Type)</b> .....	15
<b>Unexpected Antibody Detection</b> .....	16
<b>Antibody Identification</b> .....	16
<b>Compatibility Testing</b> .....	17
<b>Coagulation</b>	
<b>Prothrombin Time</b> .....	18
International Normalized Ratio (INR) .....	21
<b>Activated Partial Thromboplastin Time</b> .....	24
<b>Fibrinogen</b> .....	26
<b>Urinalysis</b>	
<b>Urinalysis Dipstick</b> .....	27
Specific Gravity .....	27
pH .....	28
Protein .....	29
Glucose .....	30
Ketones .....	31
Bilirubin .....	32
Urobilinogen .....	33
Blood or Hemoglobin .....	34
Leukocyte Esterase .....	35
Nitrite .....	36
Microalbumin (Dipstick Only) .....	37
<b>Urine hCG</b> .....	37
<b>Microbiology</b>	
<b>Miscellaneous Cultures</b> .....	38
<b>Antimicrobial Susceptibility Testing</b> .....	40
<b>Parasitology (PA Specimens)</b> .....	42
<b>Parasitology (FP Specimens)</b> .....	44

# Table of Contents (cont'd)

## Immunology

<b>Antinuclear Antibody</b> .....	49
Qualitative.....	49
Semi-Quantitative.....	49
<b>Anti-dsDNA</b> .....	51
<b>Anti-RNP</b> .....	51
<b>Anti-RNP/Sm</b> .....	52
<b>Anti-SSA</b> .....	52
<b>Anti-SSB</b> .....	53
<b>Anti-SSA/SSB</b> .....	53
<b>Anti-Sm</b> .....	54
<b>Rubella</b> .....	55
Qualitative.....	55
Quantitative.....	56
<b>Syphilis Serology</b> .....	57
VDRL Slide.....	57
VDRL Slide (Titer).....	58
MHA-TP.....	60
TPA.....	61
RPR.....	62
RPR (Titer).....	63
<b>Viral Markers</b> .....	66
Anti-HBc (IgM).....	66
Anti-HBc (Total/IgG).....	67
Anti-HIV.....	68
Anti-HAV (IgM).....	70
Anti-HAV (Total/IgG).....	71
HBeAg.....	72
Anti-HBs.....	73
HBsAg.....	74
Anti-HCV.....	75
<b>Toxoplasma gondii</b> .....	76
Qualitative (IgG).....	76
Quantitative (IgG).....	76
Qualitative (IgM).....	77
Quantitative (IgM).....	77
<b>Cytomegalovirus (CMV)</b> .....	78
Qualitative (IgG).....	78
Quantitative (IgG).....	78
Qualitative (IgM).....	79
Quantitative (IgM).....	79
<b>Chemistry</b>	
<b>Glycohemoglobin (GH Specimens)</b> .....	79
<b>CK-MB</b> .....	80

## EVALUATION CRITERIA

The evaluation criteria used in the MLE Program is in accordance with the Clinical Laboratory Improvement Amendments of 1988 (CLIA '88) federal requirements for proficiency testing. The criteria are included below.

### Qualitative/Semi-Quantitative

For qualitative/semi-quantitative procedures, evaluation is based on participant or referee consensus. A minimum percentage of participants must receive a passing score or the challenge is not evaluated due to lack of consensus. These percentages are listed below.

Antimicrobial Susceptibility Testing	80% Consensus
Antinuclear Antibody	80% Consensus
Blood Bank	95% Consensus
Cytomegalovirus	80% Consensus
Microalbumin (Semi-Quantitative)	80% Consensus
Parasite Identification	80% Consensus
Rubella	80% Consensus
Syphilis Serology	80% Consensus
Toxoplasma	80% Consensus
Urine Dipstick	80% Consensus
Urine hCG	80% Consensus
Viral Markers	80% Consensus

### Quantitative

For quantitative procedures, a mean and standard deviation (SD) are calculated for each peer group consisting of 10 or more laboratories. Acceptable performance is established based on a target value  $\pm$  the intervals below. An explanation on how to calculate the range of acceptability based upon these limits is also provided in your MLE Program Guide on page 37 under the heading "Acceptable Ranges for Quantitative Results."

Activated Partial Thromboplastin Time	$\pm$ 15%
Automated Differential	$\pm$ 3 SD
CK-MB (U/L)	$\pm$ 3 SD
Cytomegalovirus	$\pm$ 2 SD
Fibrinogen	$\pm$ 20%
Glycohemoglobin	$\pm$ 6%
Hematocrit	$\pm$ 6%
Hemoglobin	$\pm$ 7%
International Normalized Ratio (INR)	$\pm$ 20%
Platelet Count	$\pm$ 25%
Prothrombin Time	$\pm$ 15%
Red Blood Cell Count	$\pm$ 6%
Rubella	$\pm$ 3 SD
Specific Gravity	$\pm$ 0.010
Toxoplasma	$\pm$ 2 SD
White Blood Cell Count	$\pm$ 15%

**HEMATOLOGY W/ 5-PART DIFFERENTIAL–WHITE BLOO CELL COUNT (x K/uL)**

<u><i>Instrument</i></u>	<b>Specimen CL-6</b>						<b>Specimen CL-7</b>					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	52	5.88	0.73	12.4	5.7	4.9 - 6.8	52	6.55	0.68	10.4	6.3	5.5 - 7.6
All Abbott Cell-Dyn Instruments	17	6.82	0.16	2.3	6.8	5.7 - 7.9	17	7.38	0.23	3.2	7.5	6.2 - 8.5
Abbott Cell-Dyn Ruby	14	6.83	0.17	2.5	6.9	5.8 - 7.9	14	7.39	0.23	3.2	7.5	6.2 - 8.6
Orphee Mythic 22	34	5.42	0.34	6.3	5.5	4.6 - 6.3	34	6.15	0.40	6.5	6.2	5.2 - 7.1
<b>Specimen CL-8</b>												
All Method	52	2.65	0.42	15.7	2.6	2.2 - 3.1	51	6.46	0.60	9.3	6.4	5.4 - 7.5
All Abbott Cell-Dyn Instruments	16	3.20	0.13	4.0	3.3	2.7 - 3.7	16	7.19	0.18	2.6	7.2	6.1 - 8.3
Abbott Cell-Dyn Ruby	14	3.22	0.11	3.3	3.3	2.7 - 3.8	14	7.19	0.20	2.8	7.3	6.1 - 8.3
Orphee Mythic 22	34	2.42	0.22	9.0	2.4	2.0 - 2.8	32	6.21	0.25	4.0	6.3	5.2 - 7.2
<b>Specimen CL-10</b>												
All Method	52	2.68	0.41	15.3	2.5	2.2 - 3.1						
All Abbott Cell-Dyn Instruments	17	3.19	0.11	3.3	3.2	2.7 - 3.7						
Abbott Cell-Dyn Ruby	14	3.22	0.08	2.5	3.2	2.7 - 3.8						
Orphee Mythic 22	34	2.43	0.23	9.4	2.4	2.0 - 2.8						

**HEMATOLOGY W/ 5-PART DIFFERENTIAL-RED BLOOD CELL COUNT (x M/uL)**

<u><i>Instrument</i></u>	<b>Specimen CL-6</b>						<b>Specimen CL-7</b>					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	52	6.140	0.185	3.0	6.12	5.77 - 6.51	52	4.581	0.154	3.4	4.57	4.30 - 4.86
All Abbott Cell-Dyn Instruments	18	6.293	0.135	2.1	6.27	5.91 - 6.68	18	4.706	0.106	2.2	4.70	4.42 - 4.99
Abbott Cell-Dyn Ruby	15	6.304	0.137	2.2	6.27	5.92 - 6.69	15	4.729	0.080	1.7	4.70	4.44 - 5.02
Orphee Mythic 22	32	6.053	0.122	2.0	6.06	5.68 - 6.42	32	4.509	0.109	2.4	4.51	4.23 - 4.78
<u><i>Instrument</i></u>	<b>Specimen CL-8</b>						<b>Specimen CL-9</b>					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	53	2.081	0.070	3.4	2.08	1.95 - 2.21	51	4.548	0.123	2.7	4.55	4.27 - 4.83
All Abbott Cell-Dyn Instruments	18	2.119	0.054	2.6	2.12	1.99 - 2.25	17	4.664	0.110	2.4	4.69	4.38 - 4.95
Abbott Cell-Dyn Ruby	15	2.129	0.035	1.7	2.13	2.00 - 2.26	15	4.691	0.078	1.7	4.70	4.40 - 4.98
Orphee Mythic 22	34	2.064	0.070	3.4	2.06	1.94 - 2.19	33	4.495	0.078	1.7	4.50	4.22 - 4.77
<u><i>Instrument</i></u>	<b>Specimen CL-10</b>											
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>						
All Method	53	2.080	0.060	2.9	2.08	1.95 - 2.21						
All Abbott Cell-Dyn Instruments	18	2.110	0.058	2.7	2.12	1.98 - 2.24						
Abbott Cell-Dyn Ruby	15	2.121	0.038	1.8	2.12	1.99 - 2.25						
Orphee Mythic 22	33	2.063	0.044	2.1	2.06	1.93 - 2.19						

**HEMATOLOGY W/ 5-PART DIFFERENTIAL–HEMOGLOBIN (g/dL)**

<u><i>Instrument</i></u>	<b>Specimen CL-6</b>						<b>Specimen CL-7</b>					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	53	16.03	1.40	8.7	15.4	14.9 - 17.2	53	12.23	1.06	8.7	11.7	11.3 - 13.1
All Abbott Cell-Dyn Instruments	18	17.88	0.41	2.3	18.0	16.6 - 19.2	17	13.53	0.18	1.3	13.5	12.5 - 14.5
Abbott Cell-Dyn Ruby	15	17.95	0.41	2.3	18.0	16.6 - 19.3	14	13.58	0.14	1.0	13.6	12.6 - 14.6
Orphee Mythic 22	34	15.06	0.37	2.5	15.1	14.0 - 16.2	33	11.48	0.26	2.3	11.5	10.6 - 12.3
<u><i>Instrument</i></u>	<b>Specimen CL-8</b>						<b>Specimen CL-9</b>					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	52	5.03	0.46	9.1	4.8	4.6 - 5.4	53	12.16	1.04	8.6	11.7	11.3 - 13.1
All Abbott Cell-Dyn Instruments	18	5.62	0.15	2.6	5.6	5.2 - 6.1	17	13.40	0.35	2.6	13.4	12.4 - 14.4
Abbott Cell-Dyn Ruby	15	5.60	0.15	2.7	5.6	5.2 - 6.0	15	13.43	0.36	2.6	13.4	12.4 - 14.4
Orphee Mythic 22	33	4.71	0.15	3.1	4.7	4.3 - 5.1	34	11.46	0.27	2.3	11.5	10.6 - 12.3
<u><i>Instrument</i></u>	<b>Specimen CL-10</b>											
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>						
All Method	53	5.02	0.46	9.2	4.8	4.6 - 5.4						
All Abbott Cell-Dyn Instruments	18	5.62	0.15	2.6	5.6	5.2 - 6.1						
Abbott Cell-Dyn Ruby	15	5.59	0.14	2.5	5.6	5.1 - 6.0						
Orphee Mythic 22	33	4.73	0.12	2.5	4.7	4.4 - 5.1						

**HEMATOLOGY W/ 5-PART DIFFERENTIAL–HEMATOCRIT (percent)**

<u><i>Instrument</i></u>	<b>Specimen CL-6</b>						<b>Specimen CL-7</b>					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	52	51.61	2.06	4.0	51.8	48.5 - 54.8	51	38.57	1.31	3.4	38.5	36.2 - 40.9
All Abbott Cell-Dyn Instruments	18	51.07	2.97	5.8	50.8	48.0 - 54.2	18	38.29	1.79	4.7	38.0	35.9 - 40.6
Abbott Cell-Dyn Ruby	15	50.25	1.79	3.6	50.5	47.2 - 53.3	15	37.85	1.17	3.1	38.0	35.5 - 40.2
Orphee Mythic 22	34	52.11	1.76	3.4	52.4	48.9 - 55.3	33	38.89	1.24	3.2	39.0	36.5 - 41.3
<b>Specimen CL-8</b>												
All Method	53	15.98	0.84	5.3	15.9	15.0 - 17.0	51	38.53	1.29	3.3	38.5	36.2 - 40.9
All Abbott Cell-Dyn Instruments	17	15.28	0.51	3.4	15.3	14.3 - 16.2	17	37.74	1.22	3.2	37.5	35.4 - 40.0
Abbott Cell-Dyn Ruby	15	15.30	0.47	3.0	15.3	14.3 - 16.3	15	37.61	1.09	2.9	37.5	35.3 - 39.9
Orphee Mythic 22	34	16.26	0.72	4.4	16.3	15.2 - 17.3	33	38.95	1.15	3.0	39.0	36.6 - 41.3
<b>Specimen CL-10</b>												
All Method	52	15.96	0.80	5.0	15.9	15.0 - 17.0						
All Abbott Cell-Dyn Instruments	17	15.22	0.49	3.2	15.1	14.3 - 16.2						
Abbott Cell-Dyn Ruby	15	15.26	0.47	3.1	15.1	14.3 - 16.2						
Orphee Mythic 22	33	16.29	0.58	3.5	16.4	15.3 - 17.3						



**HEMATOLOGY W/ 5-PART DIFFERENTIAL-PLATELET COUNT (x K/uL)**

<u><i>Instrument</i></u>	<b>Specimen CL-6</b>						<b>Specimen CL-7</b>					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	52	345.5	22.2	6.4	349	259 - 432	51	267.0	20.0	7.5	262	200 - 334
All Abbott Cell-Dyn Instruments	18	357.5	26.8	7.5	353	268 - 447	17	251.0	8.9	3.6	248	188 - 314
Abbott Cell-Dyn Ruby	15	360.1	26.4	7.3	354	270 - 451	15	251.7	9.1	3.6	248	188 - 315
Orphee Mythic 22	34	343.3	20.7	6.0	347	257 - 430	32	275.8	17.9	6.5	273	206 - 345
<u><i>Instrument</i></u>	<b>Specimen CL-8</b>						<b>Specimen CL-9</b>					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	53	94.1	14.8	15.7	98	70 - 118	53	266.1	17.8	6.7	263	199 - 333
All Abbott Cell-Dyn Instruments	18	77.2	6.2	8.0	76	57 - 97	18	252.6	12.3	4.9	252	189 - 316
Abbott Cell-Dyn Ruby	15	75.9	3.3	4.4	75	56 - 95	15	251.4	11.0	4.4	251	188 - 315
Orphee Mythic 22	34	103.6	8.3	8.1	104	77 - 130	34	273.9	15.6	5.7	273	205 - 343
<u><i>Instrument</i></u>	<b>Specimen CL-10</b>											
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>						
All Method	53	92.7	14.3	15.4	94	69 - 116						
All Abbott Cell-Dyn Instruments	18	75.8	6.4	8.4	75	56 - 95						
Abbott Cell-Dyn Ruby	15	74.3	4.4	5.9	75	55 - 93						
Orphee Mythic 22	34	102.0	7.3	7.1	103	76 - 128						

**HEMATOLOGY W/ 5-PART DIFFERENTIAL–NEUTROPHILS (percent)**

<u><i>Instrument</i></u>	<b>Specimen CL-6</b>						<b>Specimen CL-7</b>					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	51	49.99	2.20	4.4	49.8	43.3 - 56.6	51	64.89	1.70	2.6	65.0	59.7 - 70.0
All Abbott Cell-Dyn Instruments	16	52.42	1.01	1.9	52.3	49.3 - 55.5	17	66.57	1.10	1.7	66.5	63.2 - 69.9
Abbott Cell-Dyn Ruby	13	52.45	1.08	2.1	52.5	49.2 - 55.7	14	66.51	1.15	1.7	66.5	63.0 - 70.0
Orphee Mythic 22	34	48.84	1.63	3.3	49.2	43.9 - 53.8	33	64.07	1.29	2.0	64.3	60.2 - 68.0
<b>Specimen CL-8</b>												
All Method	52	44.32	1.44	3.2	44.4	40.0 - 48.7	52	64.66	1.89	2.9	64.2	58.9 - 70.4
All Abbott Cell-Dyn Instruments	17	45.48	1.14	2.5	45.3	42.0 - 49.0	17	66.65	1.39	2.1	66.3	62.4 - 70.9
Abbott Cell-Dyn Ruby	14	45.64	1.19	2.6	45.6	42.0 - 49.3	14	66.74	1.52	2.3	66.6	62.1 - 71.4
Orphee Mythic 22	34	43.73	1.22	2.8	43.8	40.0 - 47.4	34	63.66	1.22	1.9	63.8	59.9 - 67.4
<b>Specimen CL-10</b>												
All Method	52	44.09	1.66	3.8	44.1	39.1 - 49.1						
All Abbott Cell-Dyn Instruments	17	45.21	1.33	2.9	45.1	41.2 - 49.3						
Abbott Cell-Dyn Ruby	14	45.14	1.22	2.7	45.2	41.4 - 48.8						
Orphee Mythic 22	34	43.60	1.54	3.5	43.3	38.9 - 48.3						



**HEMATOLOGY W/ 5-PART DIFFERENTIAL–MONOCYTES (percent)**

<u>Instrument</u>	<b>Specimen CL-6</b>						<b>Specimen CL-7</b>					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	52	16.78	9.44	56.2	18.0	0.0 - 45.1	52	11.57	5.35	46.3	12.2	0.0 - 27.7
All Abbott Cell-Dyn Instruments	17	5.16	0.70	13.6	4.9	3.0 - 7.3	17	5.05	0.78	15.5	4.9	2.6 - 7.4
Abbott Cell-Dyn Ruby	14	5.31	0.69	13.0	5.1	3.2 - 7.4	14	4.96	0.79	16.0	4.8	2.5 - 7.4
Orphee Mythic 22	34	22.30	5.80	26.0	22.2	4.8 - 39.8	34	14.61	3.30	22.6	15.1	4.7 - 24.6
<b>Specimen CL-8</b>												
All Method	52	18.29	9.61	52.5	19.4	0.0 - 47.2	52	11.58	5.41	46.7	12.5	0.0 - 27.8
All Abbott Cell-Dyn Instruments	17	6.55	1.54	23.6	6.5	1.9 - 11.2	17	4.98	0.84	17.0	4.6	2.4 - 7.6
Abbott Cell-Dyn Ruby	14	6.08	0.97	16.0	6.3	3.1 - 9.0	14	4.91	0.89	18.1	4.6	2.2 - 7.6
Orphee Mythic 22	34	23.88	5.97	25.0	24.1	5.9 - 41.8	34	14.73	3.38	22.9	14.9	4.6 - 24.9
<b>Specimen CL-10</b>												
All Method	52	18.47	9.63	52.2	19.3	0.0 - 47.4						
All Abbott Cell-Dyn Instruments	17	6.61	0.96	14.6	6.3	3.7 - 9.6						
Abbott Cell-Dyn Ruby	14	6.55	1.00	15.2	6.3	3.5 - 9.6						
Orphee Mythic 22	34	24.12	5.90	24.5	24.9	6.4 - 41.9						





## BLOOD BANK

### ABO GROUP

<u>Specimen</u>	<u>Results</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
BB-6	Group B	41	100%	Acceptable
BB-7	Group O	41	100%	Acceptable
BB-8	Group A	41	100%	Acceptable
BB-9	Group AB	41	100%	Acceptable
BB-10	Group A	41	100%	Acceptable

### RH FACTOR (D TYPE)

<u>Specimen</u>	<u>Results</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
BB-6	Rh Positive	41	100%	Acceptable
BB-7	Rh Negative	41	100%	Acceptable
BB-8	Rh Positive	40	97.56%	Acceptable
	Rh Negative	1	2.44%	
BB-9	Rh Negative	41	100%	Acceptable
BB-10	Rh Negative	41	100%	Acceptable

## BLOOD BANK

### UNEXPECTED ANTIBODY DETECTION

<u>Specimen</u>	<u>Results</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
AB-6	Unexpected antibody detected	23	92.00%	Acceptable
	No unexpected antibody detected	2	8.00%	
AB-7	Unexpected antibody detected	23	92.00%	Acceptable
	No unexpected antibody detected	2	8.00%	
AB-8	No unexpected antibody detected	25	100%	Acceptable
AB-9	No unexpected antibody detected	25	100%	Acceptable
AB-10	Unexpected antibody detected	20	80.00%	Acceptable
	No unexpected antibody detected	5	20.00%	

### ANTIBODY IDENTIFICATION

<u>Specimen</u>	<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
AB-6	Anti-c	14	100%	Acceptable
AB-7	Anti-D	14	100%	Acceptable
AB-8	No antibody detected	14	100%	Acceptable
AB-9	No antibody detected	14	100%	Acceptable
AB-10	Anti-K	13	92.86%	Acceptable
	No antibody detected	1	7.14%	



## BLOOD BANK

### COMPATIBILITY TESTING

<u>Specimen</u>	<u>Results</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
AB-6	Not Compatible	24	92.31%	Acceptable
	Compatible	2	7.69%	
AB-7	Compatible	25	96.15%	Acceptable
	Not Compatible	1	3.85%	
AB-8	Compatible	26	100%	Acceptable
AB-9	Compatible	26	100%	Acceptable
AB-10	Compatible	26	100%	Acceptable

# Coagulation

## PROTHROMBIN TIME (seconds)

<u>Reagent/Instrument</u>	Specimen CG-6						Specimen CG-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	87	22.68	3.13	13.8	23.1	19.2 - 26.1	57	36.12	5.98	16.6	38.0	30.7 - 41.6
Dade Innovin												
Dade Behring BFT II	5	21.75	1.98	9.1	21.0	18.4 - 25.1	5	27.05	2.54	9.4	27.4	22.9 - 31.2
Sysmex CA-500/600 series	18	19.67	1.09	5.5	19.4	16.7 - 22.7	1	-	-	-	29.8	25.3 - 34.3
All Coagulation Instruments	23	19.70	0.80	4.0	19.5	16.7 - 22.7	7	27.90	2.62	9.4	27.8	23.7 - 32.1
Dade Thromborel S												
All Coagulation Instruments	5	21.35	0.35	1.7	21.4	18.1 - 24.6	1	-	-	-	29.5	25.0 - 34.0
Diag Stago STA Neoplastine CI+												
Diagnostica Stago STA Compact	5	24.80	0.14	0.6	24.8	21.0 - 28.6	5	36.60	1.98	5.4	36.6	31.1 - 42.1
Diagnostica Stago STart 4/8	5	25.60	1.93	7.5	25.4	21.7 - 29.5	5	37.05	2.24	6.0	38.1	31.4 - 42.7
RAL Clot-SP	19	25.62	1.23	4.8	25.5	21.7 - 29.5	19	38.69	2.56	6.6	38.6	32.8 - 44.5
All Coagulation Instruments	29	25.43	1.56	6.1	25.0	21.6 - 29.3	28	38.11	2.53	6.6	38.3	32.3 - 43.9
Diagnostica Stago Neoplastine CI Plus												
Diagnostica Stago STA Compact	5	27.90	0.28	1.0	27.9	23.7 - 32.1	5	38.10	0.01	0.0	38.1	32.3 - 43.9
HemosIL RecombiPlasTin 2G												
IL ACL, all models	5	23.33	0.81	3.5	23.3	19.8 - 26.9	5	32.83	5.39	16.4	30.5	27.9 - 37.8
IL TEST PT Fibrinogen												
IL ACL, all models	6	18.85	1.00	5.3	19.2	16.0 - 21.7	6	28.30	4.84	17.1	30.0	24.0 - 32.6
IL TEST PT-FIB HS PLUS												
IL ACL, all models	5	25.26	0.69	2.7	25.2	21.4 - 29.1	5	42.92	2.09	4.9	42.9	36.4 - 49.4
PH/CMS Thromboplastin-D												
All Coagulation Instruments	5	16.90	0.71	4.2	16.9	14.3 - 19.5	5	29.95	3.61	12.0	30.0	25.4 - 34.5

**PROTHROMBIN TIME (seconds)**

<u>Reagent/Instrument</u>	Specimen CG-8						Specimen CG-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	87	29.90	4.54	15.2	30.6	25.4 - 34.4	85	12.46	1.38	11.1	12.5	10.5 - 14.4
Dade Innovin												
Dade Behring BFT II	5	29.03	1.22	4.2	28.9	24.6 - 33.4	5	10.65	0.98	9.2	10.3	9.0 - 12.3
Sysmex CA-500/600 series	18	25.99	1.03	4.0	26.1	22.0 - 29.9	17	11.02	0.37	3.4	11.0	9.3 - 12.7
All Coagulation Instruments	25	26.59	1.55	5.8	26.3	22.6 - 30.6	24	11.01	0.53	4.8	11.0	9.3 - 12.7
Dade Thromborel S												
All Coagulation Instruments	5	29.05	0.92	3.2	29.1	24.6 - 33.5	1	-	-	-	12.7	10.7 - 14.7
Diag Stago STA Neoplastine CI+												
Diagnostica Stago STA Compact	5	33.95	0.64	1.9	34.0	28.8 - 39.1	5	13.30	0.01	0.0	13.3	11.3 - 15.3
Diagnostica Stago SStart 4/8	5	34.10	2.47	7.3	34.3	28.9 - 39.3	5	12.90	0.27	2.1	13.0	10.9 - 14.9
RAL Clot-SP	19	34.46	1.24	3.6	34.7	29.2 - 39.7	19	13.99	0.41	3.0	13.8	11.8 - 16.1
All Coagulation Instruments	29	34.21	2.05	6.0	34.4	29.0 - 39.4	29	13.67	0.61	4.5	13.7	11.6 - 15.8
Diagnostica Stago Neoplastine CI Plus												
Diagnostica Stago STA Compact	5	33.20	0.85	2.6	33.2	28.2 - 38.2	5	13.15	0.21	1.6	13.2	11.1 - 15.2
HemosIL RecombiPlasTin 2G												
IL ACL, all models	5	32.00	1.61	5.0	32.5	27.2 - 36.8	5	11.03	0.05	0.5	11.0	9.3 - 12.7
IL TEST PT Fibrinogen												
IL ACL, all models	6	22.15	0.75	3.4	22.2	18.8 - 25.5	6	12.20	0.15	1.3	12.2	10.3 - 14.1
IL TEST PT-FIB HS PLUS												
IL ACL, all models	5	33.08	0.91	2.7	33.0	28.1 - 38.1	5	13.20	0.23	1.8	13.1	11.2 - 15.2
PH/CMS Thromboplastin-D												
All Coagulation Instruments	5	20.40	0.28	1.4	20.4	17.3 - 23.5	5	12.15	0.92	7.6	12.2	10.3 - 14.0

**PROTHROMBIN TIME (seconds)**

**Specimen CG-10**

<u>Reagent/Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	87	13.17	1.49	11.3	13.1	11.1 - 15.2
Dade Innovin						
Dade Behring BFT II	5	11.08	0.40	3.6	11.2	9.4 - 12.8
Sysmex CA-500/600 series	18	11.44	0.38	3.3	11.3	9.7 - 13.2
All Coagulation Instruments	24	11.37	0.39	3.4	11.3	9.6 - 13.1
Dade Thromborel S						
All Coagulation Instruments	5	11.60	1.56	13.4	11.6	9.8 - 13.4
Diag Stago STA Neoplastine Cl+						
Diagnostica Stago STA Compact	5	13.85	0.21	1.5	13.9	11.7 - 16.0
Diagnostica Stago STart 4/8	5	13.48	0.78	5.8	13.9	11.4 - 15.5
RAL Clot-SP	19	14.82	0.41	2.8	14.8	12.5 - 17.1
All Coagulation Instruments	29	14.40	0.76	5.3	14.5	12.2 - 16.6
Diagnostica Stago Neoplastine Cl Plus						
Diagnostica Stago STA Compact	5	13.95	0.21	1.5	14.0	11.8 - 16.1
HemosIL RecombiPlasTin 2G						
IL ACL, all models	5	12.73	0.21	1.6	12.7	10.8 - 14.7
IL TEST PT Fibrinogen						
IL ACL, all models	6	13.32	0.69	5.2	13.1	11.3 - 15.4
IL TEST PT-FIB HS PLUS						
IL ACL, all models	5	14.50	0.34	2.3	14.5	12.3 - 16.7
PH/CMS Thromboplastin-D						
All Coagulation Instruments	5	12.65	0.49	3.9	12.7	10.7 - 14.6

**PROTHROMBIN TIME–INTERNATIONAL NORMALIZED RATIO (INR)**

<b><u>Reagent/Instrument</u></b>	<b>Specimen CG-6</b>						<b>Specimen CG-7</b>					
	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>
All Method	83	2.11	0.27	12.9	2.1	1.6 - 2.6	51	3.70	0.79	21.4	3.7	2.9 - 4.5
Dade Innovin												
Dade Behring BFT II	5	2.18	0.10	4.4	2.2	1.7 - 2.7	5	2.65	0.31	11.7	2.7	2.1 - 3.2
Sysmex CA-500/600 series	17	1.94	0.07	3.6	1.9	1.5 - 2.4	1	-	-	-	2.5	2.0 - 3.0
All Coagulation Instruments	25	1.98	0.14	7.1	2.0	1.5 - 2.4	7	2.67	0.30	11.2	2.5	2.1 - 3.3
Dade Thromborel S												
All Coagulation Instruments	5	2.00	0.14	7.1	2.0	1.6 - 2.4	1	-	-	-	3.2	2.5 - 3.9
Diag Stago STA Neoplastine CI+												
Diagnostica Stago STA Compact	5	2.05	0.21	10.3	2.1	1.6 - 2.5	5	3.25	0.78	23.9	3.3	2.6 - 3.9
Diagnostica Stago STart 4/8	5	2.63	0.15	5.8	2.6	2.1 - 3.2	5	4.10	0.46	11.2	4.2	3.2 - 5.0
RAL Clot-SP	17	2.34	0.16	6.8	2.3	1.8 - 2.9	17	3.99	0.35	8.6	4.0	3.1 - 4.8
All Coagulation Instruments	24	2.36	0.24	10.2	2.3	1.8 - 2.9	24	3.93	0.44	11.3	4.0	3.1 - 4.8
Diagnostica Stago Neoplastine CI Plus												
Diagnostica Stago STA Compact	5	2.35	0.07	3.0	2.4	1.8 - 2.9	5	3.50	0.01	0.0	3.5	2.8 - 4.2
All Coagulation Instruments	6	2.27	0.15	6.7	2.3	1.8 - 2.8	6	3.67	0.29	7.9	3.5	2.9 - 4.5
HemosIL RecombiPlasTin 2G												
IL ACL, all models	5	2.00	0.20	10.0	2.1	1.6 - 2.4	5	2.87	0.67	23.2	2.7	2.2 - 3.5
IL TEST PT Fibrinogen												
IL ACL, all models	6	2.27	0.19	8.2	2.3	1.8 - 2.8	5	5.47	0.40	7.4	5.4	4.3 - 6.6
IL TEST PT-FIB HS PLUS												
IL ACL, all models	5	2.08	0.16	7.9	2.0	1.6 - 2.5	5	3.70	0.39	10.6	3.7	2.9 - 4.5
PH/CMS Thromboplastin-D												
All Coagulation Instruments	5	2.20	0.01	0.0	2.2	1.7 - 2.7	1	-	-	-	5.3	4.2 - 6.4

**PROTHROMBIN TIME–INTERNATIONAL NORMALIZED RATIO (INR)**

<b><u>Reagent/Instrument</u></b>	<b>Specimen CG-8</b>						<b>Specimen CG-9</b>					
	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>
All Method	82	2.90	0.46	15.7	2.8	2.3 - 3.5	82	1.04	0.10	9.3	1.0	0.8 - 1.3
Dade Innovin												
Dade Behring BFT II	5	2.80	0.14	5.1	2.9	2.2 - 3.4	5	1.18	0.10	8.1	1.2	0.9 - 1.5
Sysmex CA-500/600 series	18	2.60	0.13	5.1	2.6	2.0 - 3.2	17	1.08	0.06	5.2	1.1	0.8 - 1.3
All Coagulation Instruments	25	2.63	0.15	5.8	2.6	2.1 - 3.2	24	1.09	0.07	6.6	1.1	0.8 - 1.4
Dade Thromborel S												
All Coagulation Instruments	5	2.85	0.49	17.4	2.9	2.2 - 3.5	5	0.95	0.21	22.3	1.0	0.7 - 1.2
Diag Stago STA Neoplastine CI+												
Diagnostica Stago STA Compact	5	2.90	0.42	14.6	2.9	2.3 - 3.5	5	1.00	0.01	0.0	1.0	0.8 - 1.2
Diagnostica Stago STart 4/8	5	3.83	0.15	4.0	3.8	3.0 - 4.6	5	1.07	0.06	5.4	1.1	0.8 - 1.3
RAL Clot-SP	17	3.45	0.17	5.0	3.4	2.7 - 4.2	17	1.05	0.06	6.0	1.0	0.8 - 1.3
All Coagulation Instruments	24	3.47	0.35	10.1	3.5	2.7 - 4.2	24	1.05	0.07	6.3	1.0	0.8 - 1.3
Diagnostica Stago Neoplastine CI Plus												
Diagnostica Stago STA Compact	5	2.95	0.07	2.4	3.0	2.3 - 3.6	5	0.90	0.01	0.0	0.9	0.7 - 1.1
All Coagulation Instruments	6	3.07	0.21	6.8	3.0	2.4 - 3.7	6	0.93	0.06	6.2	0.9	0.7 - 1.2
HemosIL RecombiPlasTin 2G												
IL ACL, all models	5	2.83	0.36	12.7	3.0	2.2 - 3.4	5	0.90	0.08	9.1	0.9	0.7 - 1.1
IL TEST PT Fibrinogen												
IL ACL, all models	6	3.00	0.20	6.7	3.1	2.4 - 3.6	6	1.07	0.05	4.8	1.1	0.8 - 1.3
IL TEST PT-FIB HS PLUS												
IL ACL, all models	5	2.80	0.26	9.4	2.7	2.2 - 3.4	5	0.98	0.04	4.6	1.0	0.7 - 1.2
PH/CMS Thromboplastin-D												
All Coagulation Instruments	5	3.25	0.21	6.5	3.3	2.6 - 3.9	5	1.15	0.07	6.1	1.2	0.9 - 1.4

**PROTHROMBIN TIME-INTERNATIONAL NORMALIZED RATIO (INR)**

**Specimen CG-10**

<b><u>Reagent/Instrument</u></b>	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>
All Method	83	1.12	0.09	8.3	1.1	0.8 - 1.4
Dade Innovin						
Dade Behring BFT II	5	1.23	0.05	4.1	1.2	0.9 - 1.5
Sysmex CA-500/600 series	18	1.12	0.06	5.8	1.1	0.8 - 1.4
All Coagulation Instruments	25	1.14	0.08	7.2	1.1	0.9 - 1.4
Dade Thromborel S						
All Coagulation Instruments	5	1.00	0.14	14.1	1.0	0.8 - 1.2
Diag Stago STA Neoplastine CI+						
Diagnostica Stago STA Compact	5	1.05	0.07	6.7	1.1	0.8 - 1.3
Diagnostica Stago STart 4/8	5	1.13	0.06	5.1	1.1	0.9 - 1.4
RAL Clot-SP	17	1.14	0.07	6.2	1.1	0.9 - 1.4
All Coagulation Instruments	24	1.13	0.08	6.6	1.1	0.9 - 1.4
Diagnostica Stago Neoplastine CI Plus						
Diagnostica Stago STA Compact	5	1.00	0.01	0.0	1.0	0.8 - 1.2
All Coagulation Instruments	6	1.03	0.06	5.6	1.0	0.8 - 1.3
HemosIL RecombiPlasTin 2G						
IL ACL, all models	5	1.08	0.13	11.7	1.1	0.8 - 1.3
IL TEST PT Fibrinogen						
IL ACL, all models	6	1.23	0.08	6.6	1.2	0.9 - 1.5
IL TEST PT-FIB HS PLUS						
IL ACL, all models	5	1.10	0.07	6.4	1.1	0.8 - 1.4
PH/CMS Thromboplastin-D						
All Coagulation Instruments	5	1.25	0.07	5.7	1.3	1.0 - 1.5

**ACTIVATED PARTIAL THROMBOPLASTIN (seconds)**

<b><u>Reagent/Instrument</u></b>	<b>Specimen CG-6</b>						<b>Specimen CG-7</b>					
	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>
All Method	44	37.2	4.3	11.5	38	31 - 43	31	79.7	14.9	18.7	79	67 - 92
Dade Actin FSL												
Sysmex CA-500/600 series	10	33.4	1.6	4.7	34	28 - 39	8	66.8	6.1	9.2	67	56 - 77
All Coagulation Instruments	12	33.8	1.7	4.9	34	28 - 39	10	66.8	6.1	9.1	67	56 - 77
Diagnostica Stago STA C.K. Prest												
Diagnostica Stago STA Compact	5	42.7	0.6	1.4	43	36 - 50	5	94.7	2.5	2.7	95	80 - 109
Diagnostica Stago STart 4/8	5	41.0	4.2	10.3	41	34 - 48	5	87.0	9.9	11.4	87	73 - 101
All Coagulation Instruments	10	42.0	2.3	5.6	43	35 - 49	10	91.6	6.7	7.3	94	77 - 106
Diagnostica Stago STA-PTT												
All Coagulation Instruments	5	43.0	0.8	1.9	43	36 - 50	5	94.5	8.7	9.2	95	80 - 109
Hemoliance SynthASil												
IL ACL, all models	5	38.5	3.5	9.2	39	32 - 45	1	-	-	-	71	60 - 82
HemosIL APTT-SP												
All Coagulation Instruments	5	39.0	0.1	0.0	39	33 - 45	5	81.5	12.0	14.7	82	69 - 94
IL TEST APTT-SP												
IL ACL, all models	9	38.9	1.5	3.7	38	33 - 45	9	84.7	5.5	6.5	85	71 - 98



**ACTIVATED PARTIAL THROMBOPLASTIN (seconds)**

<u>Reagent/Instrument</u>	Specimen CG-8						Specimen CG-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	44	43.3	5.1	11.8	44	36 - 50	45	29.1	5.7	19.4	30	24 - 34
Dade Actin FSL												
Sysmex CA-500/600 series	10	39.3	1.2	3.0	40	33 - 46	10	25.2	3.5	14.0	24	21 - 29
All Coagulation Instruments	12	39.6	1.2	3.1	40	33 - 46	11	24.3	0.9	3.7	24	20 - 28
Diagnostica Stago STA C.K. Prest												
Diagnostica Stago STA Compact	5	46.7	2.1	4.5	46	39 - 54	5	32.0	1.0	3.1	32	27 - 37
Diagnostica Stago STart 4/8	5	47.5	3.5	7.4	48	40 - 55	5	32.0	1.4	4.4	32	27 - 37
All Coagulation Instruments	10	47.0	2.3	5.0	46	39 - 55	10	32.0	1.0	3.1	32	27 - 37
Diagnostica Stago STA-PTT												
All Coagulation Instruments	5	51.5	2.1	4.0	52	43 - 60	5	33.0	2.9	8.9	33	28 - 38
Hemoliance SynthASil												
IL ACL, all models	5	45.0	2.8	6.3	45	38 - 52	5	42.0	2.8	6.7	42	35 - 49
HemosIL APTT-SP												
All Coagulation Instruments	5	44.0	0.1	0.0	44	37 - 51	5	38.0	4.2	11.2	38	32 - 44
IL TEST APTT-SP												
IL ACL, all models	9	45.6	1.7	3.7	46	38 - 53	9	29.9	1.8	6.1	30	25 - 35

<u>Reagent/Instrument</u>	Specimen CG-10					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	45	28.5	3.6	12.5	29	24 - 33
Dade Actin FSL						
Sysmex CA-500/600 series	10	26.0	3.4	12.9	25	22 - 30
All Coagulation Instruments	12	26.3	3.3	12.5	25	22 - 31
Diagnostica Stago STA C.K. Prest						
Diagnostica Stago STA Compact	5	31.3	0.6	1.8	31	26 - 37
Diagnostica Stago STart 4/8	5	31.5	0.7	2.2	32	26 - 37
All Coagulation Instruments	10	31.4	0.5	1.7	31	26 - 37
Diagnostica Stago STA-PTT						
All Coagulation Instruments	5	32.8	2.2	6.8	33	27 - 38
Hemoliance SynthASil						
IL ACL, all models	5	30.0	1.4	4.7	30	25 - 35
HemosIL APTT-SP						
All Coagulation Instruments	5	30.0	1.4	4.7	30	25 - 35
IL TEST APTT-SP						
IL ACL, all models	9	30.6	1.0	3.3	30	25 - 36

**FIBRINOGEN (mg/dL)**

Specimen CG-6							Specimen CG-7					
<u>Reagent/Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	13	281.8	32.9	11.7	281	225 - 339	10	89.5	11.6	12.9	94	71 - 108
Diagnostica Stago STA Fibrinogen												
Diagnostica Stago STA Compact	5	274.8	20.0	7.3	280	219 - 330	5	96.0	3.7	3.9	97	76 - 116
All Coagulation Instruments	6	274.4	17.3	6.3	273	219 - 330	6	96.8	3.7	3.8	97	77 - 117
IL Fibrinogen-C												
IL ACL, all models	5	312.3	24.0	7.7	319	249 - 375	5	83.3	17.0	20.4	82	66 - 100
Specimen CG-8							Specimen CG-9					
<u>Reagent/Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	13	286.1	35.8	12.5	289	228 - 344	13	414.2	108.9	26.3	450	331 - 497
Diagnostica Stago STA Fibrinogen												
Diagnostica Stago STA Compact	5	273.5	16.1	5.9	277	218 - 329	5	442.0	34.9	7.9	454	353 - 531
All Coagulation Instruments	6	272.4	14.2	5.2	271	217 - 327	6	444.2	30.6	6.9	453	355 - 534
IL Fibrinogen-C												
IL ACL, all models	5	324.5	7.2	2.2	327	259 - 390	5	487.8	42.3	8.7	486	390 - 586
Specimen CG-10												
<u>Reagent/Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>						
All Method	13	288.7	46.8	16.2	289	230 - 347						
Diagnostica Stago STA Fibrinogen												
Diagnostica Stago STA Compact	5	285.8	21.9	7.7	293	228 - 343						
All Coagulation Instruments	6	282.8	20.1	7.1	289	226 - 340						
IL Fibrinogen-C												
IL ACL, all models	5	339.8	27.6	8.1	342	271 - 408						

**URINALYSIS DIPSTICK–SPECIFIC GRAVITY**

**Specimen UA-2**

<b><u>Method</u></b>	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>
All Method	365	1.0179	0.0060	0.6	1.018	1.007 - 1.028
All Arkray Methods	16	1.0281	0.0034	0.3	1.030	1.018 - 1.039
All DIRUI Methods	19	1.0203	0.0032	0.3	1.020	1.010 - 1.031
All Refractive Index Methods	36	1.0259	0.0028	0.3	1.025	1.015 - 1.036
All Roche Methods	121	1.0128	0.0043	0.4	1.010	1.002 - 1.023
All Siemens Methods	36	1.0199	0.0033	0.3	1.020	1.009 - 1.030
77 Elektronika LabUMat/2	12	1.0250	0.0030	0.3	1.026	1.015 - 1.035
Acon Laboratories	10	1.0165	0.0024	0.2	1.015	1.006 - 1.027
Arkray Aution Sticks	11	1.0292	0.0019	0.2	1.030	1.019 - 1.040
Combi-Screen Test Strips	11	1.0155	0.0027	0.3	1.015	1.005 - 1.026
DIRUI H-100 / H-500 Urine Analyzer	14	1.0200	0.0028	0.3	1.020	1.010 - 1.030
Other Analyzer Method	11	1.0208	0.0058	0.6	1.020	1.010 - 1.031
Roche Chemstrips / Combur	14	1.0118	0.0031	0.3	1.010	1.001 - 1.022
Roche cobas 6500 / u 601	12	1.0249	0.0017	0.2	1.025	1.014 - 1.035
Roche cobas u 411	81	1.0122	0.0025	0.2	1.010	1.002 - 1.023
Roche Urisys	40	1.0143	0.0065	0.6	1.010	1.004 - 1.025
SD UroColor Reagent Strips	35	1.0213	0.0033	0.3	1.020	1.011 - 1.032
Siemens Clinitek Advantus	17	1.0179	0.0025	0.2	1.020	1.007 - 1.028
Siemens Clinitek Status / Status+	16	1.0213	0.0029	0.3	1.020	1.011 - 1.032
UriScan Pro/II	12	1.0183	0.0053	0.5	1.018	1.008 - 1.029
UriScan Reagent Strips	16	1.0221	0.0026	0.3	1.020	1.012 - 1.033

## URINALYSIS DIPSTICK-pH

### Specimen UA-2

### Participant Results

<u>Method</u>	<u>Labs</u>	<u>≤3.5</u>	<u>4.0</u>	<u>4.5</u>	<u>5.0</u>	<u>5.5</u>	<u>6.0</u>	<u>6.5</u>	<u>7.0</u>	<u>7.5</u>	<u>8.0</u>	<u>8.5</u>	<u>≥9.0</u>
ALL METHODS	389	-	-	-	1	-	3	5	298	79	3	-	-
77 Elektronika LabUMat/2	14	-	-	-	-	-	-	-	13	1	-	-	-
Acon Laboratories	11	-	-	-	-	-	-	-	9	1	1	-	-
Analyticon CombiScan 500	4	-	-	-	-	-	-	-	4	-	-	-	-
Arkray Aution Jet	2	-	-	-	-	-	-	-	1	1	-	-	-
Arkray Aution Sticks	14	-	-	-	-	-	-	-	12	2	-	-	-
Arkray PocketChem UA	2	-	-	-	-	-	-	-	-	2	-	-	-
Combi-Screen Test Strips	12	-	-	-	-	-	-	-	12	-	-	-	-
DIRUI H-100 / H-500 Urine Analyzer	14	-	-	-	-	-	-	3	6	5	-	-	-
DIRUI H-800 Urine Analyzer	5	-	-	-	-	-	-	-	3	2	-	-	-
HUMAN Combilyzer	2	-	-	-	-	-	-	-	2	-	-	-	-
Iris Diagnostics Aution Max AX-4280	2	-	-	-	-	-	-	-	2	-	-	-	-
Iris Diagnostics iChem Velocity Strips	5	-	-	-	-	-	-	-	4	1	-	-	-
Iris Ichem VELOCITY Urine Chemistry System	3	-	-	-	-	-	-	-	3	-	-	-	-
Other Analyzer Method	14	-	-	-	-	-	-	-	9	5	-	-	-
Other Dipstick Method	6	-	-	-	-	-	1	-	4	1	-	-	-
Plasmatec URIPATH	1	-	-	-	-	-	-	-	1	-	-	-	-
Roche Chemstrips / Combur	19	-	-	-	-	-	-	-	19	-	-	-	-
Roche cobas 6500 / u 601	11	-	-	-	-	-	1	-	10	-	-	-	-
Roche cobas u 411	79	-	-	-	-	-	-	-	79	-	-	-	-
Roche Urisys	41	-	-	-	-	-	-	-	41	-	-	-	-
SD UroColor Reagent Strips	37	-	-	-	-	-	1	-	26	10	-	-	-
Siemens Clinitek 500	1	-	-	-	-	-	-	-	1	-	-	-	-
Siemens Clinitek Advantus	20	-	-	-	-	-	-	-	5	15	-	-	-
Siemens Clinitek Atlas	2	-	-	-	-	-	-	-	-	2	-	-	-
Siemens Clinitek Status / Status+	17	-	-	-	1	-	-	-	4	12	-	-	-
Siemens Multistix Pro	1	-	-	-	-	-	-	-	-	1	-	-	-
Siemens Reagent Strips	14	-	-	-	-	-	-	-	-	14	-	-	-
Urinometer	1	-	-	-	-	-	-	-	1	-	-	-	-
UriScan Pro/II	12	-	-	-	-	-	-	2	9	1	-	-	-
UriScan Reagent Strips	16	-	-	-	-	-	-	-	14	2	-	-	-
URIT Medical Uritest Analyzers	2	-	-	-	-	-	-	-	1	-	1	-	-
URIT Medical Uritest Reagent Strips	1	-	-	-	-	-	-	-	-	1	-	-	-

**URINALYSIS DIPSTICK-PROTEIN QUALITATIVE**  
Specimen UA-2

**Participant Results**

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>	<u>10 - 20</u> <u>mg/dL</u>	<u>30 - 70</u> <u>mg/dL</u>	<u>75</u> <u>mg/dL</u>	<u>100 - 200</u> <u>mg/dL</u>	<u>≥300 - 600</u> <u>mg/dL</u>	<u>&gt;600 or ≥1000</u> <u>mg/dL</u>
ALL METHODS	391	3	-	18	125	39	1	1	11	14	170	8	1
77 Elektronika LabUMat/2	14	-	-	-	8	-	-	-	-	-	6	-	-
Acon Laboratories	11	-	-	-	5	-	-	-	1	-	4	1	-
Analyticon CombiScan 500	4	-	-	-	1	-	-	-	-	-	3	-	-
Arkray Aution Jet	3	-	-	1	2	-	-	-	-	-	-	-	-
Arkray Aution Sticks	13	-	-	-	10	1	-	-	-	-	2	-	-
Arkray PocketChem UA	2	-	-	-	2	-	-	-	-	-	-	-	-
Combi-Screen Test Strips	12	-	-	-	1	1	-	-	1	-	8	1	-
DIRUI H-100 / H-500 Urine Analyzer	15	-	-	-	6	4	-	-	-	-	2	3	-
DIRUI H-800 Urine Analyzer	4	-	-	1	1	1	-	-	-	-	1	-	-
HUMAN Combilyzer	2	-	-	-	2	-	-	-	-	-	-	-	-
Iris Diagnostics Aution Max AX-4280	2	-	-	1	-	-	-	-	-	-	1	-	-
Iris Diagnostics iChem Velocity Strips	5	-	-	1	2	-	-	-	1	-	1	-	-
Iris Ichem VELOCITY Urine Chemistry System	3	-	-	1	-	-	-	-	-	-	2	-	-
Other Analyzer Method	16	-	-	1	11	-	-	-	-	-	3	-	1
Other Dipstick Method	6	-	-	-	4	1	-	-	-	-	1	-	-
Plasmatec URIPATH	1	-	-	1	-	-	-	-	-	-	-	-	-
Roche Chemstrips / Combur	19	-	-	2	13	3	-	-	-	-	1	-	-
Roche cobas 6500 / u 601	12	-	-	-	-	1	-	-	-	-	10	1	-
Roche cobas u 411	80	-	-	-	1	-	-	-	1	11	67	-	-
Roche Urisys	40	1	-	-	1	5	-	-	-	3	30	-	-
SD UroColor Reagent Strips	37	-	-	7	19	-	-	1	5	-	4	1	-
Siemens Clinitek 500	1	-	-	-	-	-	-	-	-	-	1	-	-
Siemens Clinitek Advantus	20	-	-	-	9	-	-	-	-	-	10	1	-
Siemens Clinitek Atlas	2	-	-	-	-	-	-	-	-	-	2	-	-
Siemens Clinitek Status / Status+	17	1	-	-	2	14	-	-	-	-	-	-	-
Siemens Reagent Strips	15	-	-	-	10	3	-	-	-	-	2	-	-
Urinometer	1	-	-	1	-	-	-	-	-	-	-	-	-
UriScan Pro/II	12	1	-	-	5	-	-	-	1	-	5	-	-
UriScan Reagent Strips	16	-	-	1	9	3	1	-	1	-	1	-	-
URIT Medical Uritest Analyzers	2	-	-	-	1	-	-	-	-	-	1	-	-
URIT Medical Uritest Reagent Strips	1	-	-	-	-	1	-	-	-	-	-	-	-

## URINALYSIS DIPSTICK–GLUCOSE

### Specimen UA-2

<u>Method</u>	<u>Labs</u>	<u>Negative or Normal</u>	<u>Trace</u>	<u>(1+)</u>	<u>Participant Results</u>			<u>30 - 100 mg/dL</u>	<u>150 - 300 mg/dL</u>	<u>500 mg/dL</u>	<u>&gt;500 or ≥1000 or ≥2000 mg/dL</u>
					<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>				
ALL METHODS	390	386	-	2	-	-	-	1	1	-	-
77 Elektronika LabUMat/2	14	14	-	-	-	-	-	-	-	-	-
Acon Laboratories	11	11	-	-	-	-	-	-	-	-	-
Analyticon CombiScan 500	4	4	-	-	-	-	-	-	-	-	-
Arkray Aution Jet	3	3	-	-	-	-	-	-	-	-	-
Arkray Aution Sticks	13	12	-	1	-	-	-	-	-	-	-
Arkray PocketChem UA	2	2	-	-	-	-	-	-	-	-	-
Combi-Screen Test Strips	12	12	-	-	-	-	-	-	-	-	-
DIRUI H-100 / H-500 Urine Analyzer	15	15	-	-	-	-	-	-	-	-	-
DIRUI H-800 Urine Analyzer	4	4	-	-	-	-	-	-	-	-	-
HUMAN Combilyzer	2	2	-	-	-	-	-	-	-	-	-
Iris Diagnostics Aution Max AX-4280	2	2	-	-	-	-	-	-	-	-	-
Iris Diagnostics iChem Velocity Strips	5	5	-	-	-	-	-	-	-	-	-
Iris Ichem VELOCITY Urine Chemistry System	3	3	-	-	-	-	-	-	-	-	-
Other Analyzer Method	15	15	-	-	-	-	-	-	-	-	-
Other Dipstick Method	6	6	-	-	-	-	-	-	-	-	-
Plasmatec URIPATH	1	1	-	-	-	-	-	-	-	-	-
Roche Chemstrips / Combur	19	19	-	-	-	-	-	-	-	-	-
Roche cobas 6500 / u 601	12	12	-	-	-	-	-	-	-	-	-
Roche cobas u 411	80	79	-	-	-	-	-	-	1	-	-
Roche Urisys	40	40	-	-	-	-	-	-	-	-	-
SD UroColor Reagent Strips	37	36	-	-	-	-	-	1	-	-	-
Siemens Clinitek 500	1	1	-	-	-	-	-	-	-	-	-
Siemens Clinitek Advantus	20	20	-	-	-	-	-	-	-	-	-
Siemens Clinitek Atlas	2	2	-	-	-	-	-	-	-	-	-
Siemens Clinitek Status / Status+	17	16	-	1	-	-	-	-	-	-	-
Siemens Reagent Strips	15	15	-	-	-	-	-	-	-	-	-
Urinometer	1	1	-	-	-	-	-	-	-	-	-
UriScan Pro/II	12	12	-	-	-	-	-	-	-	-	-
UriScan Reagent Strips	16	16	-	-	-	-	-	-	-	-	-
URIT Medical Uritest Analyzers	2	2	-	-	-	-	-	-	-	-	-
URIT Medical Uritest Reagent Strips	1	1	-	-	-	-	-	-	-	-	-

## URINALYSIS DIPSTICK–KETONES

### Specimen UA-2

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>Small</u>	<u>Moderate</u>	<u>Large</u>	<u>Participant Results</u>					<u>5 - 10 mg/dL</u>	<u>15 - 25 mg/dL</u>	<u>40 - 60 mg/dL</u>	<u>80 - 100 mg/dL</u>	<u>≥150 mg/dL</u>
							<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>						
ALL METHODS	389	386	-	-	-	-	1	1	1	-	-	-	-	-	-	-
77 Elektronika LabUMat/2	14	14	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Acon Laboratories	11	11	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Analyticon CombiScan 500	4	4	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Arkray Aution Jet	3	3	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Arkray Aution Sticks	13	12	-	-	-	-	1	-	-	-	-	-	-	-	-	-
Arkray PocketChem UA	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Combi-Screen Test Strips	12	12	-	-	-	-	-	-	-	-	-	-	-	-	-	-
DIRUI H-100 / H-500 Urine Analyzer	15	15	-	-	-	-	-	-	-	-	-	-	-	-	-	-
DIRUI H-800 Urine Analyzer	4	4	-	-	-	-	-	-	-	-	-	-	-	-	-	-
HUMAN Combilyzer	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Iris Diagnostics Aution Max AX-4280	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Iris Diagnostics iChem Velocity Strips	5	5	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Iris Ichem VELOCITY Urine Chemistry System	3	3	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Other Analyzer Method	14	14	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Other Dipstick Method	6	6	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Plasmatec URIPATH	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Roche Chemstrips / Combur	18	18	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Roche cobas 6500 / u 601	12	12	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Roche cobas u 411	80	80	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Roche Urisys	40	40	-	-	-	-	-	-	-	-	-	-	-	-	-	-
SD UroColor Reagent Strips	37	36	-	-	-	-	-	1	-	-	-	-	-	-	-	-
Siemens Clinitek 500	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Advantus	20	20	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Atlas	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Status / Status+	17	16	-	-	-	-	-	-	-	1	-	-	-	-	-	-
Siemens Reagent Strips	15	15	-	-	-	-	-	-	-	-	-	-	-	-	-	-
Urinometer	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-
UriScan Pro/II	12	12	-	-	-	-	-	-	-	-	-	-	-	-	-	-
UriScan Reagent Strips	16	16	-	-	-	-	-	-	-	-	-	-	-	-	-	-
URIT Medical Uritest Analyzers	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-
URIT Medical Uritest Reagent Strips	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-

**URINALYSIS DIPSTICK–BILIRUBIN**

**Specimen UA-2**

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>Small</u>	<u>Moderate</u>	<u>Participant Results</u>					<u>0.5 - 1.0 mg/dL</u>	<u>2.0 - 4.0 mg/dL</u>	<u>6.0 - 10.0 mg/dL</u>	<u>&gt;10.0 mg/dL</u>
						<u>Large</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>				
ALL METHODS	374	372	-	-	-	-	-	-	1	-	-	1	-	-
77 Elektronika LabUMat/2	14	14	-	-	-	-	-	-	-	-	-	-	-	-
Acon Laboratories	12	12	-	-	-	-	-	-	-	-	-	-	-	-
Analyticon CombiScan 500	5	5	-	-	-	-	-	-	-	-	-	-	-	-
Arkray Aution Jet	3	3	-	-	-	-	-	-	-	-	-	-	-	-
Arkray Aution Sticks	13	13	-	-	-	-	-	-	-	-	-	-	-	-
Arkray PocketChem UA	2	2	-	-	-	-	-	-	-	-	-	-	-	-
Combi-Screen Test Strips	11	11	-	-	-	-	-	-	-	-	-	-	-	-
DIRUI H-100 / H-500 Urine Analyzer	15	15	-	-	-	-	-	-	-	-	-	-	-	-
DIRUI H-800 Urine Analyzer	4	4	-	-	-	-	-	-	-	-	-	-	-	-
HUMAN Combilyzer	2	2	-	-	-	-	-	-	-	-	-	-	-	-
Iris Diagnostics Aution Max AX-4280	2	2	-	-	-	-	-	-	-	-	-	-	-	-
Iris Diagnostics iChem Velocity Strips	6	6	-	-	-	-	-	-	-	-	-	-	-	-
Iris Ichem VELOCITY Urine Chemistry System	2	2	-	-	-	-	-	-	-	-	-	-	-	-
Other Analyzer Method	15	15	-	-	-	-	-	-	-	-	-	-	-	-
Other Dipstick Method	6	6	-	-	-	-	-	-	-	-	-	-	-	-
Plasmatec URIPATH	1	1	-	-	-	-	-	-	-	-	-	-	-	-
Roche Chemstrips / Combur	16	16	-	-	-	-	-	-	-	-	-	-	-	-
Roche cobas 6500 / u 601	12	12	-	-	-	-	-	-	-	-	-	-	-	-
Roche cobas u 411	80	80	-	-	-	-	-	-	-	-	-	-	-	-
Roche Urisys	40	40	-	-	-	-	-	-	-	-	-	-	-	-
SD UroColor Reagent Strips	37	36	-	-	-	-	-	-	1	-	-	-	-	-
Siemens Clinitek 500	1	1	-	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Advantus	20	19	-	-	-	-	-	-	-	-	-	1	-	-
Siemens Clinitek Atlas	2	2	-	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Status / Status+	16	16	-	-	-	-	-	-	-	-	-	-	-	-
Siemens Reagent Strips	4	4	-	-	-	-	-	-	-	-	-	-	-	-
UriScan Pro/II	12	12	-	-	-	-	-	-	-	-	-	-	-	-
UriScan Reagent Strips	16	16	-	-	-	-	-	-	-	-	-	-	-	-
URIT Medical Uritest Analyzers	2	2	-	-	-	-	-	-	-	-	-	-	-	-
URIT Medical Uritest Reagent Strips	1	1	-	-	-	-	-	-	-	-	-	-	-	-



## URINALYSIS DIPSTICK–UROBILINOGEN

### Specimen UA-2

<u>Method</u>	<u>Labs</u>	<i>Participant Results</i>				
		<u>Normal or 0.0 - 0.2 mg/dL or &lt;3.2 µmol/L</u>	<u>1.0 or &lt;2.0 mg/dL or 16 µmol/L</u>	<u>2.0/3.0 mg/dL or 34 or 35 µmol/L</u>	<u>4.0 or 4.0/6.0 mg/dL or 70 µmol/L</u>	<u>&gt;=8.0 or 12.0 mg/dL or &gt;=140 or 200 µmol/L</u>
ALL METHODS	374	373	-	-	1	-
77 Elektronika LabUMat/2	14	14	-	-	-	-
Acon Laboratories	11	11	-	-	-	-
Analyticon CombiScan 500	4	4	-	-	-	-
Arkray Aution Jet	3	3	-	-	-	-
Arkray Aution Sticks	13	12	-	-	1	-
Arkray PocketChem UA	2	2	-	-	-	-
Combi-Screen Test Strips	12	12	-	-	-	-
DIRUI H-100 / H-500 Urine Analyzer	15	15	-	-	-	-
DIRUI H-800 Urine Analyzer	4	4	-	-	-	-
HUMAN Combilyzer	2	2	-	-	-	-
Iris Diagnostics Aution Max AX-4280	2	2	-	-	-	-
Iris Diagnostics iChem Velocity Strips	6	6	-	-	-	-
Iris Ichem VELOCITY Urine Chemistry System	2	2	-	-	-	-
Other Analyzer Method	15	15	-	-	-	-
Other Dipstick Method	6	6	-	-	-	-
Plasmatec URIPATH	1	1	-	-	-	-
Roche Chemstrips / Combur	16	16	-	-	-	-
Roche cobas 6500 / u 601	12	12	-	-	-	-
Roche cobas u 411	79	79	-	-	-	-
Roche Urisys	40	40	-	-	-	-
SD UroColor Reagent Strips	37	37	-	-	-	-
Siemens Clinitek 500	1	1	-	-	-	-
Siemens Clinitek Advantus	20	20	-	-	-	-
Siemens Clinitek Atlas	2	2	-	-	-	-
Siemens Clinitek Status / Status+	16	16	-	-	-	-
Siemens Reagent Strips	4	4	-	-	-	-
Urinometer	1	1	-	-	-	-
UriScan Pro/II	12	12	-	-	-	-
UriScan Reagent Strips	16	16	-	-	-	-
URIT Medical Uritest Analyzers	2	2	-	-	-	-
URIT Medical Uritest Reagent Strips	1	1	-	-	-	-

**URINALYSIS DIPSTICK–BLOOD/HEMOGLOBIN**

Specimen UA-2

**Participant Results**

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>Small</u>	<u>Moderate</u>	<u>Large</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>	<u>(5+)</u>	<u>5 - 25</u> <u>Ery/<math>\mu</math>L</u>	<u>50 -</u> <u>100</u> <u>Ery/<math>\mu</math>L</u>	<u>250</u> <u>Ery/<math>\mu</math>L</u>	<u><math>\pm</math>0.03</u> <u>mg/dL</u>	<u>0.06</u> <u>-</u> <u>0.10</u> <u>mg/</u> <u>dL</u>	<u>0.2 -</u> <u>0.5</u> <u>mg/</u> <u>dL</u>	<u><math>\geq</math>1.0</u> <u>mg/</u> <u>dL</u>
ALL METHODS	387	3	1	-	-	5	-	8	167	40	1	-	3	155	-	-	1	3
77 Elektronika LabUMat/2	14	-	-	-	-	-	-	-	11	1	-	-	-	2	-	-	-	-
Acon Laboratories	10	-	-	-	-	-	-	-	7	3	-	-	-	-	-	-	-	-
Analyticon CombiScan 500	4	-	-	-	-	-	-	-	2	-	-	-	-	2	-	-	-	-
Arkray Aution Jet	3	-	-	-	-	-	-	-	3	-	-	-	-	-	-	-	-	-
Arkray Aution Sticks	13	-	-	-	-	-	-	2	11	-	-	-	-	-	-	-	-	-
Arkray PocketChem UA	2	-	-	-	-	-	-	1	1	-	-	-	-	-	-	-	-	-
Combi-Screen Test Strips	12	-	-	-	-	1	-	-	5	-	-	-	-	6	-	-	-	-
DIRUI H-100 / H-500 Urine Analyzer	15	-	-	-	-	-	-	-	12	1	-	-	1	1	-	-	-	-
DIRUI H-800 Urine Analyzer	4	-	-	-	-	-	-	-	3	-	-	-	-	1	-	-	-	-
HUMAN Combilyzer	2	-	-	-	-	-	-	-	1	1	-	-	-	-	-	-	-	-
Iris Diagnostics Aution Max AX-4280	2	-	-	-	-	-	-	1	-	-	-	-	-	-	-	-	1	-
Iris Diagnostics iChem Velocity Strips	6	1	-	-	-	-	-	-	4	-	-	-	-	-	-	-	-	1
Iris Ichem VELOCITY Urine Chemistry System	2	-	-	-	-	-	-	-	-	1	-	-	-	-	-	-	-	1
Other Analyzer Method	15	-	1	-	-	-	-	3	10	-	-	-	-	1	-	-	-	-
Other Dipstick Method	6	-	-	-	-	-	-	-	4	1	-	-	-	1	-	-	-	-
Plasmatec URIPATH	1	-	-	-	-	-	-	-	1	-	-	-	-	-	-	-	-	-
Roche Chemstrips / Combur	17	-	-	-	-	-	-	-	2	13	-	-	-	2	-	-	-	-
Roche cobas 6500 / u 601	12	-	-	-	-	-	-	-	-	-	1	-	-	11	-	-	-	-
Roche cobas u 411	80	-	-	-	-	-	-	-	-	2	-	-	1	77	-	-	-	-
Roche Mditron Junior/II	1	-	-	-	-	-	-	-	-	1	-	-	-	-	-	-	-	-
Roche Urisys	40	1	-	-	-	-	-	-	1	5	-	-	-	33	-	-	-	-
SD UroColor Reagent Strips	37	1	-	-	-	-	-	-	25	8	-	-	-	3	-	-	-	-
Siemens Clinitek 500	1	-	-	-	-	-	-	-	-	-	-	-	-	1	-	-	-	-
Siemens Clinitek Advantus	19	-	-	-	-	1	-	-	13	-	-	-	1	4	-	-	-	-
Siemens Clinitek Atlas	2	-	-	-	-	-	-	-	-	-	-	-	-	2	-	-	-	-
Siemens Clinitek Status / Status+	16	-	-	-	-	2	-	-	14	-	-	-	-	-	-	-	-	-
Siemens Reagent Strips	16	-	-	-	-	1	-	-	14	1	-	-	-	-	-	-	-	-
Urinometer	1	-	-	-	-	-	-	-	1	-	-	-	-	-	-	-	-	-
UriScan Pro/II	12	-	-	-	-	-	-	1	4	-	-	-	-	7	-	-	-	-
UriScan Reagent Strips	16	-	-	-	-	-	-	-	13	2	-	-	-	-	-	-	-	1
URIT Medical Uritest Analyzers	2	-	-	-	-	-	-	-	2	-	-	-	-	-	-	-	-	-
URIT Medical Uritest Reagent Strips	1	-	-	-	-	-	-	-	1	-	-	-	-	-	-	-	-	-

**URINALYSIS DIPSTICK–LEUKOCYTE ESTERASE**

**Specimen UA-2**

<u>Method</u>	<i>Participant Results</i>												
	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>Small</u>	<u>Moderate</u>	<u>Large</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>	<u>15 or 25 µL</u>	<u>75 or 100 µL</u>	<u>250 or 500 µL</u>
ALL METHODS	374	364	4	-	-	-	2	2	-	1	-	-	1
77 Elektronika LabUMat/2	14	14	-	-	-	-	-	-	-	-	-	-	-
Acon Laboratories	11	11	-	-	-	-	-	-	-	-	-	-	-
Analyticon CombiScan 500	4	4	-	-	-	-	-	-	-	-	-	-	-
Arkray Aution Jet	3	3	-	-	-	-	-	-	-	-	-	-	-
Arkray Aution Sticks	12	11	1	-	-	-	-	-	-	-	-	-	-
Arkray PocketChem UA	2	2	-	-	-	-	-	-	-	-	-	-	-
Combi-Screen Test Strips	13	13	-	-	-	-	-	-	-	-	-	-	-
DIRUI H-100 / H-500 Urine Analyzer	15	15	-	-	-	-	-	-	-	-	-	-	-
DIRUI H-800 Urine Analyzer	4	4	-	-	-	-	-	-	-	-	-	-	-
HUMAN Combilyzer	2	1	1	-	-	-	-	-	-	-	-	-	-
Iris Diagnostics Aution Max AX-4280	2	2	-	-	-	-	-	-	-	-	-	-	-
Iris Diagnostics iChem Velocity Strips	6	6	-	-	-	-	-	-	-	-	-	-	-
Iris Ichem VELOCITY Urine Chemistry System	2	2	-	-	-	-	-	-	-	-	-	-	-
Other Analyzer Method	15	15	-	-	-	-	-	-	-	-	-	-	-
Other Dipstick Method	6	6	-	-	-	-	-	-	-	-	-	-	-
Plasmatec URIPATH	1	1	-	-	-	-	-	-	-	-	-	-	-
Roche Chemstrips / Combur	17	17	-	-	-	-	-	-	-	-	-	-	-
Roche cobas 6500 / u 601	12	11	-	-	-	-	-	-	-	-	-	-	1
Roche cobas u 411	81	80	-	-	-	-	-	-	-	1	-	-	-
Roche Urisys	40	40	-	-	-	-	-	-	-	-	-	-	-
SD UroColor Reagent Strips	35	30	2	-	-	-	1	2	-	-	-	-	-
Siemens Clinitek 500	1	1	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Advantus	20	20	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Atlas	2	2	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Status / Status+	16	15	-	-	-	-	1	-	-	-	-	-	-
Siemens Reagent Strips	4	4	-	-	-	-	-	-	-	-	-	-	-
Urinometer	1	1	-	-	-	-	-	-	-	-	-	-	-
UriScan Pro/II	11	11	-	-	-	-	-	-	-	-	-	-	-
UriScan Reagent Strips	16	16	-	-	-	-	-	-	-	-	-	-	-
URIT Medical Uritest Analyzers	2	2	-	-	-	-	-	-	-	-	-	-	-
URIT Medical Uritest Reagent Strips	1	1	-	-	-	-	-	-	-	-	-	-	-

## URINALYSIS DIPSTICK–NITRITE

### Specimen UA-2

#### Participant Results

<u>Method</u>	<u>Labs</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	377	376	1
77 Elektronika LabUMat/2	14	14	-
Acon Laboratories	11	11	-
Analyticon CombiScan 500	4	4	-
Arkray Aution Jet	3	3	-
Arkray Aution Sticks	13	13	-
Arkray PocketChem UA	2	2	-
Combi-Screen Test Strips	12	12	-
DIRUI H-100 / H-500 Urine Analyzer	15	15	-
DIRUI H-800 Urine Analyzer	4	4	-
HUMAN Combilyzer	2	2	-
Iris Diagnostics Aution Max AX-4280	2	2	-
Iris Diagnostics iChem Velocity Strips	6	6	-
Iris Ichem VELOCITY Urine Chemistry System	2	2	-
Other Analyzer Method	15	15	-
Other Dipstick Method	6	6	-
Plasmatec URIPATH	1	1	-
Roche Chemstrips / Combur	18	18	-
Roche cobas 6500 / u 601	12	12	-
Roche cobas u 411	80	80	-
Roche Urisys	40	39	1
SD UroColor Reagent Strips	37	37	-
Siemens Clinitek 500	1	1	-
Siemens Clinitek Advantus	20	20	-
Siemens Clinitek Atlas	2	2	-
Siemens Clinitek Status / Status+	16	16	-
Siemens Reagent Strips	4	4	-
Urinometer	1	1	-
UriScan Pro/II	12	12	-
UriScan Reagent Strips	16	16	-
URIT Medical Uritest Analyzers	2	2	-
URIT Medical Uritest Reagent Strips	1	1	-

**URINALYSIS –MICROALBUMIN (dipstick only)**

**Specimen UA-2**

*Participant Results*

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>10 mg/L</u>	<u>20 mg/L</u>	<u>30 mg/L</u>	<u>50 mg/L</u>	<u>80 mg/L</u>	<u>100 mg/L</u>	<u>150 mg/L</u>	<u>+(4 - 8 mg/dL)</u>	<u>++ (&gt;8 mg/dL)</u>
ALL METHODS	23	5	1	-	-	-	2	3	7	1	4
Arkray Aution Sticks	1	1	-	-	-	-	-	-	-	-	-
Beckman Coulter ICON microALB	1	-	-	-	-	-	1	-	-	-	-
Other Analyzer Method	6	1	-	-	-	-	-	1	1	1	2
Roche cobas 6000 / c 501	3	-	-	-	-	-	-	-	3	-	-
Roche cobas u 411	1	1	-	-	-	-	-	-	-	-	-
Roche Micral - 1 minute	4	1	-	-	-	-	1	1	1	-	-
Siemens Clinitek Microalbumin	1	-	-	-	-	-	-	1	-	-	-
UriScan Reagent Strips	2	1	1	-	-	-	-	-	-	-	-

**URINALYSIS –URINE hCG**

**Specimen UA-2**

*Participant Results*

<u>Method</u>	<u>Labs</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	112	111	1
Acon Laboratories	5	5	-
Alere Clearview hCG Cassette	7	7	-
Alere hCG Cassette	12	12	-
Biotron 1-Step	1	1	-
Medline hCG Test Strip	1	1	-
Quidel QuickVue One-Step Combo	19	19	-
Quidel QuickVue One-Step Urine	1	1	-
SD Bioline hCG	16	16	-
Siemens Clinitek Status / Status+	10	9	1
Stanbio QuStick	1	1	-

## MISCELLANEOUS CULTURES

### Specimen BA-4 – Blood Culture

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Staphylococcus haemolyticus	83	82.18%	Acceptable
Staphylococcus sp.	8	7.92%	Acceptable
Staph – coagulase neg.	4	3.96%	Acceptable
Staphylococcus epidermidis	3	2.97%	
Staphylococcus aureus	2	1.98%	
Staphylococcus hominis	1	0.99%	

Organism(s) present: *Staphylococcus haemolyticus*.

### Specimen BA-5 – Stool Culture

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Escherichia coli	80	47.34%	Acceptable
Shigella sp.	63	37.28%	Acceptable
Shigella flexneri	21	12.43%	Acceptable
Normal flora	1	0.59%	Acceptable
Gram positive cocci	3	1.78%	
No enteric pathogens isolated	1	0.59%	

Organism(s) present: *Shigella flexneri* and *Escherichia coli*.

## MISCELLANEOUS CULTURES

### Specimen BA-6 – Wound Culture

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Streptococcus salivarius	72	41.14%	Acceptable
Eikenella corrodens	42	24.00%	Acceptable
Streptococcus alpha-hemolytic	13	7.43%	Acceptable
Eikenella sp.	9	5.14%	Acceptable
Gram negative bacilli	4	2.29%	Acceptable
Gram positive cocci	4	2.29%	Acceptable
Streptococcus non-hemolytic	3	1.71%	Acceptable
Gram negative coccobacilli	3	1.71%	Acceptable
Streptococcus viridans group	2	1.14%	Acceptable
Escherichia coli	4	2.29%	
Streptococcus mitis	3	1.71%	
Aeromonas sp.	2	1.14%	
Haemophilus sp.	2	1.14%	
Strep – beta hemo, not Grp A	1	0.57%	
Staphylococcus epidermidis	1	0.57%	
Serratia sp.	1	0.57%	
Gram positive bacilli	1	0.57%	
Corynebacterium sp.	1	0.57%	
Enterococcus sp.	1	0.57%	
Anaerobe present – no ID	1	0.57%	
Anaerobic cultures not performance	1	0.57%	
Pasteurella sp.	1	0.57%	
Prevotella sp.	1	0.57%	
Fusobacterium nucleatum	1	0.57%	
Streptococcus intermedius	1	0.57%	

Organism(s) present: *Streptococcus salivarius* and *Eikenella corrodens*.

**ANTIMICROBIAL SUSCEPTIBILIY TESTING**

<b>Specimen UC-6, CC-6 (SUS-6)</b>	<b>-----Disk Diffusion-----</b>				<b>-----MIC-----</b>				<b><u>Acceptable (%)</u></b>
	<b><i>Interpretative category data</i></b>				<b><i>Interpretative category data</i></b>				
	<b><u>Labs</u></b>	<b><u>S</u></b>	<b><u>I</u></b>	<b><u>R</u></b>	<b><u>Labs</u></b>	<b><u>S</u></b>	<b><u>I</u></b>	<b><u>R</u></b>	
Amikacin	38	37	-	1	119	118	-	1	98.15%
Amoxicillin/Clavulanate	34	31	1	2	49	48	-	1	95.18%
Ampicillin	19	-	-	19	63	2	1	60	96.47%
Ampicillin/Sulbactam	22	19	1	2	80	78	-	2	95.19%
Aztreonam	18	18	-	-	33	33	-	-	100.00%
Carbenicillin	-	-	-	-	1	-	-	1	Inappropriate drug <sup>1</sup>
Cefaclor	4	4	-	-	1	1	-	-	100.00%
Cefamandole	-	-	-	-	1	1	-	-	Ungraded <sup>2</sup>
Cefazolin	15	15	-	-	48	48	-	-	100.00%
Cefepime	28	28	-	-	115	112	1	2	97.96%
Cefixime	19	19	-	-	8	8	-	-	100.00%
Cefoperazone	9	9	-	-	1	1	-	-	100.00%
Cefotaxime	30	30	-	-	55	55	-	-	100.00%
Cefotetan	-	-	-	-	2	2	-	-	100.00%
Cefoxitin	8	8	-	-	40	40	-	-	100.00%
Cefpodoxime	5	5	-	-	3	3	-	-	100.00%
Ceftazidime	31	31	-	-	103	103	-	-	100.00%
Ceftizoxime	-	-	-	-	3	3	-	-	100.00%
Ceftriaxone	33	33	-	-	94	92	1	1	98.47%
Cefuroxime	34	33	1	-	75	73	-	2	97.27%
Cephalexin	4	4	-	-	3	3	-	-	Inappropriate drug <sup>1</sup>
Cephalothin	3	3	-	-	36	36	-	-	Inappropriate drug <sup>1</sup>
Ciprofloxacin	42	42	-	-	138	137	-	1	99.46%
Colistin	3	2	-	1	12	11	-	1	Inappropriate drug <sup>1</sup>
Daptomycin	-	-	-	-	1	1	-	-	Inappropriate drug <sup>1</sup>
Doripenem	2	2	-	-	4	4	-	-	100.00%
Doxycycline	1	1	-	-	2	1	-	1	Ungraded <sup>3</sup>
Ertapenem	16	16	-	-	92	90	1	1	98.20%
Fosfomycin	7	7	-	-	17	13	-	4	Inappropriate drug <sup>1</sup>
Gatifloxacin	-	-	-	-	1	1	-	-	Ungraded <sup>2</sup>

<sup>1</sup> This is an inappropriate drug for organism and/or source.

<sup>2</sup> This is an ungraded challenge due to lack of comparison group.

<sup>3</sup> This is an ungraded challenge due to less than 80% participant consensus.



ANTIMICROBIAL SUSCEPTIBILITY TESTING (cont'd)

<u>Antimicrobial</u>	-----Disk Diffusion-----				-----MIC-----				<u>Acceptable (%)</u>
	<u>Interpretative category data</u>				<u>Interpretative category data</u>				
	<u>Labs</u>	<u>S</u>	<u>I</u>	<u>R</u>	<u>Labs</u>	<u>S</u>	<u>I</u>	<u>R</u>	
Gentamicin	34	32	1	1	124	123	-	1	98.16%
Imipenem	28	28	-	-	78	78	-	-	100.00%
Kanamycin	-	-	-	-	1	1	-	-	Ungraded <sup>2</sup>
Levofloxacin	23	23	-	-	62	61	-	1	98.88%
Linezolid	-	-	-	-	1	1	-	-	Inappropriate drug <sup>1</sup>
Lomefloxacin	-	-	-	-	1	1	-	-	Ungraded <sup>2</sup>
Meropenem	25	25	-	-	99	98	1	-	99.20%
Methicillin	-	-	-	-	1	1	-	-	Inappropriate drug <sup>1</sup>
Minocycline	2	2	-	-	1	1	-	-	100.00%
Moxifloxacin	3	3	-	-	1	1	-	-	Inappropriate drug <sup>1</sup>
Nalidixic Acid	7	7	-	-	5	4	-	1	91.67%
Netilmicin	9	9	-	-	3	3	-	-	100.00%
Nitrofurantoin	32	28	3	1	93	83	10	-	89.06%
Norfloxacin	17	17	-	-	49	49	-	-	100.00%
Ofloxacin	14	14	-	-	4	3	-	1	94.44%
Oxacillin	-	-	-	-	1	1	-	-	Inappropriate drug <sup>1</sup>
Penicillin	-	-	-	-	2	1	-	1	Inappropriate drug <sup>1</sup>
Piperacillin	3	3	-	-	7	7	-	-	100.00%
Piperacillin/Tazobactam	24	24	-	-	80	79	-	1	99.06%
Quinupristin/Dalfopristin	-	-	-	-	1	1	-	-	Inappropriate drug <sup>1</sup>
Rifampin	-	-	-	-	2	2	-	-	Inappropriate drug <sup>1</sup>
Sulfisoxazole	-	-	-	-	1	1	-	-	Ungraded <sup>2</sup>
Teicoplanin	-	-	-	-	1	1	-	-	Inappropriate drug <sup>1</sup>
Tetracycline	6	6	-	-	11	10	-	1	94.12%
Ticarcillin	-	-	-	-	2	2	-	-	Inappropriate drug <sup>1</sup>
Ticarcillin/Clavulanate	1	1	-	-	5	5	-	-	100.00%
Tigecycline	-	-	-	-	5	4	1	-	Inappropriate drug <sup>1</sup>
Tobramycin	4	4	-	-	23	23	-	-	100.00%
Trimethoprim	-	-	-	-	3	3	-	-	100.00%
Trimethoprim/Sulfamethoxazole	32	32	-	-	119	118	1	-	99.35%
Vancomycin	-	-	-	-	2	1	-	1	Inappropriate drug <sup>1</sup>

Organism(s) present: *Klebsiella pneumoniae*.

NOTE: Please be aware that CLSI issues annual editions of M100, the standards used by all proficiency testing programs for grading of susceptibilities. Drugs considered appropriate may change significantly with subsequent editions. The current edition of the CLSI M100 document is accessible online at CLSI.org under Standards>Free Resources.

## PARASITOLOGY (PA Specimens)

### Specimen PA-6

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
No parasite seen	4	57.14%	Unacceptable
Entamoeba coli	1	14.29%	Unacceptable
Hookworm	1	14.29%	Unacceptable
Trichinella spiralis	1	14.29%	Unacceptable

Parasite(s) present: *Dientamoeba fragilis*. This specimen is graded to US statistics.

### Specimen PA-7

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
No parasite seen	4	40.00%	Acceptable
Dientamoeba fragilis	1	10.00%	
Entamoeba histolytica	1	10.00%	
Giardia lamblia	1	10.00%	
Hookworm	1	10.00%	
Hymenolepis nana eggs	1	10.00%	
Protozoan seen but no ID	1	10.00%	

Parasite(s) present: No parasite present. This specimen is graded to US statistics.

### Specimen PA-8

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Hymenolepis nana eggs	21	87.50%	Acceptable
Hymenolepis diminuta eggs	2	8.33%	
Ascaris lumbricoides eggs	1	4.17%	

Parasite(s) present: *Hymenolepis nana* eggs.

**PARASITOLOGY (PA Specimens) cont'd**

**Specimen PA-9**

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Entamoeba histolytica	19	76.00%	Acceptable
Protozoan seen but no ID	1	4.00%	Acceptable
Other parasite seen but no ID	1	4.00%	Acceptable
Entamoeba coli	2	4.00%	
Iodamoeba buetschlii	1	4.00%	
No parasite seen	1	4.00%	

Parasite(s) present: *Entamoeba histolytica*.

**Specimen PA-10**

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Strongyloides stercoralis larva	21	87.50%	Acceptable
Parasite larva seen but no ID	1	4.17%	Acceptable
No parasite seen	1	4.17%	
Ascaris lumbricoides eggs	1	4.17%	

Parasite(s) present: *Strongyloides stercoralis* larva.

## PARASITOLOGY (FP Specimens)

### Specimen FP-6

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Schistosoma sp. egg	150	60.48%	Acceptable
Schistosoma mansoni eggs	14	5.65%	Acceptable
Schistosoma haematobium eggs	38	15.32%	
No parasite seen	35	14.11%	
Balantidium coli	2	0.81%	
Chilomastix mesnili	1	0.40%	
Entamoeba coli	1	0.40%	
Entamoeba histolytica	1	0.40%	
Fasciola hepatica eggs	1	0.40%	
Giardia lamblia	1	0.40%	
Paragonimus westermani eggs	1	0.40%	
Trypanosoma cruzi	1	0.40%	
Ascaris lumbricoides eggs	1	0.40%	
Schistosoma japonicum eggs	1	0.40%	

Parasite(s) present: *Schistosoma mansoni* eggs. Specimen FP-6 is graded by 84% referee consensus.

## PARASITOLOGY (FP Specimens)

### Specimen FP-7

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Clonorchis sinensis	140	55.34%	Acceptable
Parasite egg seen but no ID	4	1.58%	Acceptable
Other parasite seen but no ID	1	0.40%	Acceptable
No parasite seen	55	21.74%	
Giardia lamblia	10	3.95%	
Fasciola hepatica eggs	7	2.77%	
Taenia sp. eggs	7	2.77%	
Blastocystis hominis	4	1.58%	
Endolimax nana	3	1.19%	
Entamoeba coli	3	1.19%	
Enterobius vermicularis eggs	3	1.19%	
Ascaris lumbricoides eggs	3	1.19%	
Paragonimus westermani eggs	3	1.19%	
Parasite larva seen but no ID	2	0.79%	
Trichuris trichiura eggs	2	0.79%	
Schistosoma sp. eggs	1	0.40%	
Schistosoma haematobium eggs	1	0.40%	
Strongyloides stercoralis larvae	1	0.40%	
Balantidium coli	1	0.40%	
Hookworm	1	0.40%	
Hymenolepis diminuta eggs	1	0.40%	

Parasite(s) present: *Clonorchis sinensis*. Specimen FP-7 is graded by 87% referee consensus.

**PARASITOLOGY (FP Specimens) cont'd**

**Specimen FP-8**

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Balantidium coli	210	85.71%	Acceptable
No parasite seen	14	5.71%	
Diphyllobothrium latum	3	1.22%	
Paragonimus westermani eggs	2	0.82%	
Schistosoma japonicum eggs	2	0.82%	
Blastocystis hominis	2	0.82%	
Ascaris lumbricoides eggs	2	0.82%	
Clonorchis sinensis	2	0.82%	
Hookworm	1	0.41%	
Hymenolepis diminuta eggs	1	0.41%	
Hymenolepis nana eggs	1	0.41%	
Strongyloides stercoralis larvae	1	0.41%	
Taenia sp. eggs	1	0.41%	
Parasite egg seen but no ID	1	0.41%	
Nonpath, protozoan present	1	0.41%	
Entamoeba coli	1	0.41%	

Parasite(s) present: *Balantidium coli*.

## PARASITOLOGY (FP Specimens) cont'd

### Specimen FP-9

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Diphyllobothrium latum	187	73.05%	Acceptable
Parasite egg seen but no ID	3	1.17%	Acceptable
Fasciola hepatica eggs	22	8.59%	
Paragonimus westermani eggs	12	4.69%	
Endolimax nana	8	3.13%	
Ascaris lumbricoides eggs	5	1.95%	
Hookworm	5	1.95%	
Entamoeba histolytica	4	1.56%	
Trichostrongylus sp. eggs	2	0.78%	
No parasite seen	2	0.78%	
Cyclospora cayetanensis	1	0.39%	
Nonpath, protozoan present	1	0.39%	
Balantidium coli	1	0.39%	
Iodamoeba buetschlii	1	0.39%	
Clonorchis sinensis	1	0.39%	
Trypanosoma brucei sp.	1	0.39%	

Parasite(s) present: *Diphyllobothrium latum*. This challenge was graded by 82% referee consensus

## PARASITOLOGY (FP Specimens) cont'd

### Specimen FP-10

<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Trypanosoma sp.	120	50.85%	Acceptable
Trypanosoma brucei sp.	34	14.41%	Acceptable
Trypanosoma brucei gambiense	12	14.41%	Acceptable
Trypanosoma cruzi	63	26.69%	
Trypanosoma brucei sp.	34	14.41%	
Trypanosoma brucei rhodesiense	5	2.12%	
Leishmania sp.	1	0.42%	
Plasmodium falciparum	1	0.42%	

Parasite(s) present: *Trypanosoma brucei gambiense*. This challenge was graded by 88% referee consensus.



**Antinuclear Antibody (ANA) - Qualitative**

<u>Method</u>	Specimen AE-6		Specimen AE-7		Specimen AE-8	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	1	17	14	4	1	17
Bio-Rad	-	1	1	-	-	1
BioSystems	1	-	1	-	1	-
Immuno Concepts	-	3	3	-	-	3
INOVA Diagnostics	-	6	4	2	-	6
Kallestad	-	1	1	-	-	1
Zeus	-	1	1	-	-	1

<u>Method</u>	Specimen AE-9		Specimen AE-10	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	16	2	17	1
Bio-Rad	1	-	1	-
BioSystems	1	-	1	-
Immuno Concepts	3	-	3	-
INOVA Diagnostics	5	1	5	1
Kallestad	1	-	1	-
Zeus	1	-	1	-

**Antinuclear Antibody (ANA)—Semi-Quantitative (Titer)**

<u>Specimen/Method</u>	<u>N/A</u> <u>(Neg)</u>	<u>8/</u> <u>10</u>	<u>16/</u> <u>20</u>	<u>32/</u> <u>40</u>	<u>64/</u> <u>80</u>	<u>128/</u> <u>160</u>	<u>256/</u> <u>320</u>	<u>512/</u> <u>640</u>	<u>&gt;640</u>	<u>1024/</u> <u>1280</u>	<u>2048/</u> <u>2560</u>	<u>≥2560</u>
------------------------	----------------------------	------------------------	-------------------------	-------------------------	-------------------------	---------------------------	---------------------------	---------------------------	----------------	-----------------------------	-----------------------------	--------------

**Specimen AE-6**

ALL METHODS	11	-	-	-	-	-	-	-	-	-	-	-
Immuno Concepts	2	-	-	-	-	-	-	-	-	-	-	-
INOVA Diagnostics	4	-	-	-	-	-	-	-	-	-	-	-
Kallestad	1	-	-	-	-	-	-	-	-	-	-	-
Zeus	1	-	-	-	-	-	-	-	-	-	-	-

**Antinuclear Antibody (ANA)—Semi-Quantitative (Titer)**

<u>Specimen/Method</u>	<u>N/A</u> <u>(Neg)</u>	<u>8/</u> <u>10</u>	<u>16/</u> <u>20</u>	<u>32/</u> <u>40</u>	<u>64/</u> <u>80</u>	<u>128/</u> <u>160</u>	<u>256/</u> <u>320</u>	<u>512/</u> <u>640</u>	<u>&gt;640</u>	<u>1024/</u> <u>1280</u>	<u>2048/</u> <u>2560</u>	<u>≥2560</u>
------------------------	----------------------------	------------------------	-------------------------	-------------------------	-------------------------	---------------------------	---------------------------	---------------------------	----------------	-----------------------------	-----------------------------	--------------

**Specimen AE-7**

ALL METHODS	1	-	-	-	4	-	1	1	-	3	1	-
Immuno Concepts	-	-	-	-	2	-	-	-	-	-	-	-
INOVA Diagnostics	1	-	-	-	1	-	1	-	-	1	-	-
Kallestad	-	-	-	-	-	-	-	-	-	-	1	-
Zeus	-	-	-	-	-	-	-	-	-	1	-	-

**Specimen AE-8**

ALL METHODS	11	-	-	-	-	-	-	-	-	-	-	-
Immuno Concepts	2	-	-	-	-	-	-	-	-	-	-	-
INOVA Diagnostics	4	-	-	-	-	-	-	-	-	-	-	-
Kallestad	1	-	-	-	-	-	-	-	-	-	-	-
Zeus	1	-	-	-	-	-	-	-	-	-	-	-

**Specimen AE-9**

ALL METHODS	-	-	-	-	-	1	3	2	1	4	-	-
Immuno Concepts	-	-	-	-	-	-	2	-	-	-	-	-
INOVA Diagnostics	-	-	-	-	-	-	1	1	1	1	-	-
Kallestad	-	-	-	-	-	-	-	1	-	-	-	-
Zeus	-	-	-	-	-	-	-	-	-	1	-	-

**Specimen AE-10**

ALL METHODS	1	-	-	-	-	-	2	2	-	5	1	-
Immuno Concepts	-	-	-	-	-	-	-	-	-	2	-	-
INOVA Diagnostics	1	-	-	-	-	-	1	-	-	1	1	-
Kallestad	-	-	-	-	-	-	-	-	-	1	-	-
Zeus	-	-	-	-	-	-	-	-	-	1	-	-

## Anti-dsDNA

<u>Method</u>	Specimen AE-6		Specimen AE-7		Specimen AE-8	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	-	16	-	15	-	16
BioSystems	-	1	-	1	-	1
INOVA Diagnostics	-	7	-	6	-	7
Kallestad	-	1	-	1	-	1
Zeus	-	1	-	1	-	1

<u>Method</u>	Specimen AE-9		Specimen AE-10	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	1	14	-	15
BioSystems	-	1	-	1
INOVA Diagnostics	1	5	-	6
Kallestad	-	1	-	1
Zeus	-	1	-	1

## Anti-RNP

<u>Method</u>	Specimen AE-6		Specimen AE-7		Specimen AE-8	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	-	12	-	12	-	12
INOVA Diagnostics	-	7	-	7	-	7
Kallestad	-	1	-	1	-	1

<u>Method</u>	Specimen AE-9		Specimen AE-10	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	10	2	1	11
INOVA Diagnostics	6	1	1	6
Kallestad	1	-	-	1

## Anti-RNP/Sm

<u>Method</u>	Specimen AE-6		Specimen AE-7		Specimen AE-8	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	-	7	1	6	-	7
INOVA Diagnostics	-	1	-	1	-	1

<u>Method</u>	Specimen AE-9		Specimen AE-10	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	5	2	-	7
INOVA Diagnostics	-	1	-	1

## Anti-SSA

<u>Method</u>	Specimen AE-6		Specimen AE-7		Specimen AE-8	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	-	17	-	17	-	17
INOVA Diagnostics	-	8	-	8	-	8
Kallestad	-	1	-	1	-	1

<u>Method</u>	Specimen AE-9		Specimen AE-10	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	1	16	16	1
INOVA Diagnostics	1	7	7	1
Kallestad	-	1	1	-

**Anti-SSB**

<u>Method</u>	Specimen AE-6		Specimen AE-7		Specimen AE-8	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	-	17	-	17	-	17
INOVA Diagnostics	-	8	-	8	-	8
Kallestad	-	1	-	1	-	1

<u>Method</u>	Specimen AE-9		Specimen AE-10	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	1	16	16	1
INOVA Diagnostics	1	7	7	1
Kallestad	-	1	1	-

**Anti-SSA/SSB**

<u>Method</u>	Specimen AE-6		Specimen AE-7		Specimen AE-8	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	-	1	-	1	-	1
INOVA Diagnostics	-	1	-	1	-	1

<u>Method</u>	Specimen AE-9		Specimen AE-10	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	-	1	1	-
INOVA Diagnostics	-	1	1	-

**Anti-Sm**

<u>Method</u>	Specimen AE-6		Specimen AE-7		Specimen AE-8	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	-	17	-	17	-	17
INOVA Diagnostics	-	8	-	8	-	8
Kallestad	-	1	-	1	-	1

<u>Method</u>	Specimen AE-9		Specimen AE-10	
	<u>Positive</u>	<u>Negative</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	-	17	-	17
INOVA Diagnostics	-	8	-	8
Kallestad	-	1	-	1

**Rubella—Qualitative**

<b><u>Method</u></b>	<b>Specimen RU-6</b>		<b>Specimen RU-7</b>		<b>Specimen RU-8</b>	
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>
ALL METHODS	15	1	-	16	15	1
Abbott Architect	10	1	-	11	10	1
Roche cobas 6000 / e 601	1	-	-	1	1	-
Roche cobas e 411	1	-	-	1	1	-
Siemens ADVIA Centaur	2	-	-	2	2	-
VITROS ECI	1	-	-	1	1	-

<b><u>Method</u></b>	<b>Specimen RU-9</b>		<b>Specimen RU-10</b>	
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>
ALL METHODS	15	1	-	16
Abbott Architect	10	1	-	11
Roche cobas 6000 / e 601	1	-	-	1
Roche cobas e 411	1	-	-	1
Siemens ADVIA Centaur	2	-	-	2
VITROS ECI	1	-	-	1

**Rubella—Quantitative (IU/mL)**

<u>Specimen/Method</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
<b>Specimen RU-6</b>						
All Method	21	59.04	33.75	57.2	37.6	0.0 - 160.3
Abbott Architect	13	35.57	2.79	7.8	35.6	27.2 - 44.0
<b>Specimen RU-7</b>						
All Method	23	0.11	0.16	139.4	0.0	0.0 - 0.6
Abbott Architect	12	0.00	0.01	0.0	0.0	0.0 - 0.1
<b>Specimen RU-8</b>						
All Method	21	59.43	32.78	55.2	38.5	0.0 - 157.8
Abbott Architect	13	36.50	2.30	6.3	35.8	29.5 - 43.5
<b>Specimen RU-9</b>						
All Method	21	45.00	27.64	61.4	27.2	0.0 - 128.0
Abbott Architect	13	26.18	1.17	4.5	26.1	22.6 - 29.7
<b>Specimen RU-10</b>						
All Method	22	0.07	0.11	154.1	0.0	0.0 - 0.5
Abbott Architect	13	0.00	0.01	0.0	0.0	0.0 - 0.1



**Syphilis Serology—Qualitative: VDRL Slide**

<b><u>Method</u></b>	<b>Specimen SY-6</b>			<b>Specimen SY-7</b>			<b>Specimen SY-8</b>		
	<b><u>Reactive</u></b>	<b><u>Weakly Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Weakly Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Weakly Reactive</u></b>	<b><u>Non-Reactive</u></b>
ALL METHODS	44	1	2	45	-	2	47	-	-
Abbott Architect	2	-	-	2	-	-	2	-	-
Acon Laboratories	2	-	-	2	-	-	2	-	-
Human	1	-	-	1	-	-	1	-	-
Omega Diagnostics	3	-	-	3	-	-	3	-	-
Plasmatec	1	-	-	1	-	-	1	-	-
SPINREACT	5	-	-	5	-	-	5	-	-
Standard Diagnostics	1	-	-	1	-	-	1	-	-
Wiener Lab	25	-	1	25	-	1	26	-	-

<b><u>Method</u></b>	<b>Specimen SY-9</b>			<b>Specimen SY-10</b>		
	<b><u>Reactive</u></b>	<b><u>Weakly Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Weakly Reactive</u></b>	<b><u>Non-Reactive</u></b>
ALL METHODS	-	-	47	-	-	47
Abbott Architect	-	-	2	-	-	2
Acon Laboratories	-	-	2	-	-	2
Human	-	-	1	-	-	1
Omega Diagnostics	-	-	3	-	-	3
Plasmatec	-	-	1	-	-	1
SPINREACT	-	-	5	-	-	5
Standard Diagnostics	-	-	1	-	-	1
Wiener Lab	-	-	26	-	-	26

**Syphilis Serology—Semi-Quantitative: VDRL Slide Titer**

<u>Specimen/Method</u>	<u>N/A (Neg)</u>	<u>0 dils</u>	<u>1 dil</u>	<u>2 dils</u>	<u>4 dils</u>	<u>8 dils</u>	<u>16 dils</u>	<u>32 dils</u>	<u>&gt;32 dils</u>
<b>Specimen SY-6</b>									
ALL METHODS	1	1	8	19	7	1	2	-	-
Omega Diagnostics	-	-	1	1	-	-	-	-	-
Plasmatec	-	-	-	1	-	-	-	-	-
SPINREACT	-	-	-	1	-	-	-	-	-
Wiener Lab	1	1	5	14	7	-	1	-	-
<b>Specimen SY-7</b>									
ALL METHODS	1	1	6	22	5	2	1	-	1
Omega Diagnostics	-	-	1	1	-	-	-	-	-
Plasmatec	-	-	-	1	-	-	-	-	-
SPINREACT	-	-	1	-	-	-	-	-	-
Wiener Lab	1	1	3	17	4	2	-	-	1
<b>Specimen SY-8</b>									
ALL METHODS	-	-	1	8	17	7	5	1	-
Omega Diagnostics	-	-	-	1	-	1	-	-	-
Plasmatec	-	-	-	-	1	-	-	-	-
SPINREACT	-	-	-	1	-	-	-	-	-
Wiener Lab	-	-	1	5	14	5	4	-	-

**Syphilis Serology—Semi-Quantitative: VDRL Slide Titer**

<b><u>Specimen/Method</u></b>	<b><u>N/A (Neg)</u></b>	<b><u>0 dils</u></b>	<b><u>1 dil</u></b>	<b><u>2 dils</u></b>	<b><u>4 dils</u></b>	<b><u>8 dils</u></b>	<b><u>16 dils</u></b>	<b><u>32 dils</u></b>	<b><u>&gt;32 dils</u></b>
<b>Specimen SY-9</b>									
ALL METHODS	39	-	-	-	-	-	-	-	-
Omega Diagnostics	2	-	-	-	-	-	-	-	-
Plasmatec	1	-	-	-	-	-	-	-	-
SPINREACT	1	-	-	-	-	-	-	-	-
Wiener Lab	29	-	-	-	-	-	-	-	-
<b>Specimen SY-10</b>									
ALL METHODS	39	-	-	-	-	-	-	-	-
Omega Diagnostics	2	-	-	-	-	-	-	-	-
Plasmatec	1	-	-	-	-	-	-	-	-
SPINREACT	1	-	-	-	-	-	-	-	-
Wiener Lab	29	-	-	-	-	-	-	-	-

**Syphilis Serology—Qualitative: MHA-TP**

<b><u>Method</u></b>	<b>Specimen SY-6</b>		<b>Specimen SY-7</b>		<b>Specimen SY-8</b>	
	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>
ALL METHODS	17	-	17	-	16	1
Abbott Architect	3	-	3	-	2	1
Biokit	1	-	1	-	1	-
bioMerieux	1	-	1	-	1	-
DiaSorin	1	-	1	-	1	-
Human	1	-	1	-	1	-
Plasmatec	5	-	5	-	5	-
Serodia	2	-	2	-	2	-
SPINREACT	1	-	1	-	1	-

  

	<b>Specimen SY-9</b>		<b>Specimen SY-10</b>	
	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>
ALL METHODS	-	17	-	17
Abbott Architect	-	3	-	3
Biokit	-	1	-	1
bioMerieux	-	1	-	1
DiaSorin	-	1	-	1
Human	-	1	-	1
Plasmatec	-	5	-	5
Serodia	-	2	-	2
SPINREACT	-	1	-	1

**Syphilis Serology—Qualitative: *Treponema pallidum* Antibodies**

<b><u>Method</u></b>	<b>Specimen SY-6</b>		<b>Specimen SY-7</b>		<b>Specimen SY-8</b>	
	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>
ALL METHODS	47	1	47	1	48	-
Abbott Architect	10	-	10	-	10	-
bioMerieux	1	-	1	-	1	-
Human	3	-	3	-	3	-
Plasmatec	5	-	5	-	5	-
Roche cobas 6000 / c 501	2	-	2	-	2	-
Roche cobas e 411	1	-	1	-	1	-
Serodia	10	-	10	-	10	-
SPINREACT	1	-	1	-	1	-
Standard Diagnostics	4	-	4	-	4	-

  

	<b>Specimen SY-9</b>		<b>Specimen SY-10</b>	
	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>
ALL METHODS	-	48	-	48
Abbott Architect	-	10	-	10
bioMerieux	-	1	-	1
Human	-	3	-	3
Plasmatec	-	5	-	5
Roche cobas 6000 / c 501	-	2	-	2
Roche cobas e 411	-	1	-	1
Serodia	-	10	-	10
SPINREACT	-	1	-	1
Standard Diagnostics	-	4	-	4

**Syphilis Serology—Qualitative: RPR**

<b><u>Method</u></b>	<b>Specimen SY-6</b>		<b>Specimen SY-7</b>		<b>Specimen SY-8</b>	
	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>
ALL METHODS	74	6	74	6	80	-
Abbott Architect	1	-	1	-	1	-
ASI	1	-	1	-	1	-
Becton Dickinson	1	-	1	-	1	-
bioMerieux	4	1	4	1	5	-
BioSystems	14	-	14	-	14	-
Human	6	-	6	-	6	-
Omega Diagnostics	10	2	10	2	12	-
Plasmatec	14	1	14	1	15	-
Pulse Scientific	1	-	1	-	1	-
SPINREACT	16	1	16	1	17	-
Standard Diagnostics	1	-	1	-	1	-

	<b>Specimen SY-9</b>		<b>Specimen SY-10</b>	
	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>	<b><u>Reactive</u></b>	<b><u>Non-Reactive</u></b>
ALL METHODS	-	80	-	80
Abbott Architect	-	1	-	1
ASI	-	1	-	1
Becton Dickinson	-	1	-	1
bioMerieux	-	5	-	5
BioSystems	-	14	-	14
Human	-	6	-	6
Omega Diagnostics	-	12	-	12
Plasmatec	-	15	-	15
Pulse Scientific	-	1	-	1
SPINREACT	-	17	-	17
Standard Diagnostics	-	1	-	1

**Syphilis Serology—Semi-Quantitative: RPR (Titer)**

<b><u>Specimen/Method</u></b>	<b><u>N/A (Neg)</u></b>	<b><u>1</u></b>	<b><u>2</u></b>	<b><u>4</u></b>	<b><u>8</u></b>	<b><u>16</u></b>	<b><u>32</u></b>	<b><u>64</u></b>	<b><u>&gt;64</u></b>
<b>Specimen SY-6</b>									
ALL METHODS	3	15	37	4	2	-	1	-	-
Becton Dickinson	-	-	1	-	-	-	-	-	-
bioMerieux	-	2	1	1	-	-	-	-	-
BioSystems	-	2	8	1	-	-	-	-	-
Human	-	1	4	-	1	-	-	-	-
Omega Diagnostics	2	2	5	-	-	-	1	-	-
Plasmatec	-	2	7	1	-	-	-	-	-
Pulse Scientific	-	-	-	1	-	-	-	-	-
SPINREACT	1	3	10	-	1	-	-	-	-
<b>Specimen SY-7</b>									
ALL METHODS	3	12	36	7	3	-	1	-	-
Becton Dickinson	-	-	1	-	-	-	-	-	-
bioMerieux	-	3	1	-	-	-	-	-	-
BioSystems	-	1	10	-	-	-	-	-	-
Human	-	1	4	1	-	-	-	-	-
Omega Diagnostics	2	1	5	1	-	-	1	-	-
Plasmatec	-	2	5	2	1	-	-	-	-
Pulse Scientific	-	-	-	1	-	-	-	-	-
SPINREACT	1	3	7	2	2	-	-	-	-

**Syphilis Serology—Semi-Quantitative: RPR (Titer) cont'd**

<b><u>Specimen/Method</u></b>	<b><u>N/A (Neg)</u></b>	<b><u>1</u></b>	<b><u>2</u></b>	<b><u>4</u></b>	<b><u>8</u></b>	<b><u>16</u></b>	<b><u>32</u></b>	<b><u>64</u></b>	<b><u>&gt;64</u></b>
<b>Specimen SY-8</b>									
ALL METHODS	1	1	11	34	10	4	-	1	-
Becton Dickinson	-	-	-	-	1	-	-	-	-
bioMerieux	-	-	2	1	-	1	-	-	-
BioSystems	-	-	1	7	3	-	-	-	-
Human	-	-	1	2	2	1	-	-	-
Omega Diagnostics	1	-	1	6	2	-	-	-	-
Plasmatec	-	-	1	6	1	1	-	1	-
Pulse Scientific	-	-	-	1	-	-	-	-	-
SPINREACT	-	1	4	8	1	1	-	-	-

**Specimen SY-9**

ALL METHODS	61	1	-	-	-	-	-	-	-
Becton Dickinson	1	-	-	-	-	-	-	-	-
bioMerieux	4	-	-	-	-	-	-	-	-
BioSystems	11	-	-	-	-	-	-	-	-
Human	6	-	-	-	-	-	-	-	-
Omega Diagnostics	10	-	-	-	-	-	-	-	-
Plasmatec	10	-	-	-	-	-	-	-	-
Pulse Scientific	1	-	-	-	-	-	-	-	-
SPINREACT	14	1	-	-	-	-	-	-	-
	61	1	-	-	-	-	-	-	-



Syphilis Serology—Semi-Quantitative: RPR (Titer) cont'd

<u>Specimen/Method</u>	<u>N/A</u> <u>(Neg)</u>	<u>1</u>	<u>2</u>	<u>4</u>	<u>8</u>	<u>16</u>	<u>32</u>	<u>64</u>	<u>&gt;64</u>
<b>Specimen SY-10</b>									
ALL METHODS	61	1	-	-	-	-	-	-	-
Becton Dickinson	1	-	-	-	-	-	-	-	-
bioMerieux	4	-	-	-	-	-	-	-	-
BioSystems	11	-	-	-	-	-	-	-	-
Human	6	-	-	-	-	-	-	-	-
Omega Diagnostics	10	-	-	-	-	-	-	-	-
Plasmatec	10	-	-	-	-	-	-	-	-
Pulse Scientific	1	-	-	-	-	-	-	-	-
SPINREACT	14	1	-	-	-	-	-	-	-

## Viral Markers – Anti-HBc (IgM)

<u>Method</u>	<b>Specimen VM-6</b>			<b>Specimen VM-7</b>			<b>Specimen VM-8</b>		
	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>
ALL METHODS	2	48	-	-	50	-	-	50	-
Abbott Architect	1	24	-	-	25	-	-	25	-
Beckman ACCESS / 2 / Dxl	-	1	-	-	1	-	-	1	-
bioMerieux Vidas, Mini Vidas	-	3	-	-	3	-	-	3	-
Roche cobas 6000 / e 601	-	11	-	-	11	-	-	11	-
Roche cobas e 411	-	2	-	-	2	-	-	2	-
Roche Modular Analytics	-	1	-	-	1	-	-	1	-
Siemens ADVIA Centaur	1	2	-	-	3	-	-	3	-
VITROS 3600/4600/5600	-	3	-	-	3	-	-	3	-

  

<u>Method</u>	<b>Specimen VM-9</b>			<b>Specimen VM-10</b>		
	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>
ALL METHODS	-	50	-	-	50	-
Abbott Architect	-	25	-	-	25	-
Beckman ACCESS / 2 / Dxl	-	1	-	-	1	-
bioMerieux Vidas, Mini Vidas	-	3	-	-	3	-
Roche cobas 6000 / e 601	-	11	-	-	11	-
Roche cobas e 411	-	2	-	-	2	-
Roche Modular Analytics	-	1	-	-	1	-
Siemens ADVIA Centaur	-	3	-	-	3	-
VITROS 3600/4600/5600	-	3	-	-	3	-

**Viral Markers – Anti-HBc (Total / IgG)**

<b><u>Method</u></b>	<b>Specimen VM-6</b>			<b>Specimen VM-7</b>			<b>Specimen VM-8</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	70	2	-	2	70	-	2	70	-
Abbott Architect	41	-	-	-	41	-	-	41	-
Beckman ACCESS / 2 / Dxl	2	-	-	-	2	-	-	2	-
bioMerieux Vidas, Mini Vidas	2	-	-	-	2	-	-	2	-
Roche cobas 6000 / e 601	11	2	-	2	11	-	2	11	-
Roche cobas e 411	4	-	-	-	4	-	-	4	-
Roche Modular Analytics	1	-	-	-	1	-	-	1	-
Siemens ADVIA Centaur	3	-	-	-	3	-	-	3	-
VITROS 3600/4600/5600	3	-	-	-	3	-	-	3	-
VITROS Eci	1	-	-	-	1	-	-	1	-

<b><u>Method</u></b>	<b>Specimen VM-9</b>			<b>Specimen VM-10</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	2	70	-	1	71	-
Abbott Architect	-	41	-	-	41	-
Beckman ACCESS / 2 / Dxl	-	2	-	-	2	-
bioMerieux Vidas, Mini Vidas	-	2	-	-	2	-
Roche cobas 6000 / e 601	2	11	-	1	12	-
Roche cobas e 411	-	4	-	-	4	-
Roche Modular Analytics	-	1	-	-	1	-
Siemens ADVIA Centaur	-	3	-	-	3	-
VITROS 3600/4600/5600	-	3	-	-	3	-
VITROS Eci	-	1	-	-	1	-

**Viral Markers – Anti-HIV**

<b><u>Method</u></b>	<b>Specimen VM-6</b>			<b>Specimen VM-7</b>			<b>Specimen VM-8</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	1	166	-	2	165	-	139	26	2
Abbott Architect	-	71	-	1	70	-	67	3	1
Acon Laboratories	-	1	-	-	1	-	1	-	-
Alere Clearview HIV1/2 STAT-PAK	-	1	-	-	1	-	-	1	-
Alere Determine HIV - moderate	-	4	-	-	4	-	1	3	-
Alere Determine HIV - waived	-	1	-	-	1	-	1	-	-
Beckman ACCESS / 2 / Dxl bioMerieux Vidas, Mini Vidas	-	4	-	-	4	-	4	-	-
DiaSorin	-	3	-	-	3	-	2	1	-
Human	-	1	-	-	1	-	1	-	-
Roche cobas 6000 / e 601	-	1	-	-	1	-	-	1	-
Roche cobas e 411	1	31	-	-	32	-	30	2	-
Roche Elecsys 1010 / 2010	-	15	-	1	14	-	15	-	-
Roche Elecsys 1010 / 2010	-	1	-	-	1	-	1	-	-
Roche Modular Analytics	-	3	-	-	3	-	3	-	-
Siemens ADVIA Centaur	-	6	-	-	6	-	6	-	-

**Viral Markers – Anti-HIV- cont'd**

<b><u>Method</u></b>	<b>Specimen VM-9</b>			<b>Specimen VM-10</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	1	166	-	163	4	-
Abbott Architect	-	71	-	70	1	-
Acon Laboratories	-	1	-	1	-	-
Alere Clearview HIV1/2 STAT-PAK	-	1	-	1	-	-
Alere Determine HIV - moderate	-	4	-	4	-	-
Alere Determine HIV - waived	-	1	-	1	-	-
Beckman ACCESS / 2 / DxI	-	4	-	4	-	-
bioMerieux Vidas, Mini Vidas	-	3	-	3	-	-
DiaSorin	-	1	-	1	-	-
Human	-	1	-	1	-	-
Roche cobas 6000 / e 601	1	31	-	29	3	-
Roche cobas e 411	-	15	-	15	-	-
Roche Elecsys 1010 / 2010	-	1	-	1	-	-
Roche Modular Analytics	-	3	-	3	-	-
Siemens ADVIA Centaur	-	6	-	6	-	-

**Viral Markers – Anti-HAV (IgM)**

<u>Method</u>	<b>Specimen VM-6</b>			<b>Specimen VM-7</b>			<b>Specimen VM-8</b>		
	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>
ALL METHODS	-	66	-	-	66	-	65	1	-
Abbott Architect	-	33	-	-	33	-	33	-	-
Beckman ACCESS / 2 / Dxl	-	1	-	-	1	-	1	-	-
bioMerieux Vidas, Mini Vidas	-	5	-	-	5	-	5	-	-
Roche cobas 6000 / e 601	-	15	-	-	15	-	14	1	-
Roche cobas e 411	-	2	-	-	2	-	2	-	-
Roche Modular Analytics	-	1	-	-	1	-	1	-	-
Siemens ADVIA Centaur	-	4	-	-	4	-	4	-	-
Standard Diagnostics	-	2	-	-	2	-	2	-	-
VITROS 3600/4600/5600	-	1	-	-	1	-	1	-	-
VITROS Eci	-	1	-	-	1	-	1	-	-

<u>Method</u>	<b>Specimen VM-9</b>			<b>Specimen VM-10</b>		
	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>
ALL METHODS	-	66	-	-	66	-
Abbott Architect	-	33	-	-	33	-
Beckman ACCESS / 2 / Dxl	-	1	-	-	1	-
bioMerieux Vidas, Mini Vidas	-	5	-	-	5	-
Roche cobas 6000 / e 601	-	15	-	-	15	-
Roche cobas e 411	-	2	-	-	2	-
Roche Modular Analytics	-	1	-	-	1	-
Siemens ADVIA Centaur	-	4	-	-	4	-
Standard Diagnostics	-	2	-	-	2	-
VITROS 3600/4600/5600	-	1	-	-	1	-
VITROS Eci	-	1	-	-	1	-

**Viral Markers – Anti-HAV (Total/IgG)**

<b><u>Method</u></b>	<b>Specimen VM-6</b>			<b>Specimen VM-7</b>			<b>Specimen VM-8</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	2	55	-	52	4	1	55	1	1
Abbott Architect	-	27	-	26	-	1	26	-	1
bioMerieux Vidas, Mini Vidas	-	4	-	4	-	-	4	-	-
Roche cobas 6000 / e 601	2	11	-	12	1	-	13	-	-
Roche cobas e 411	-	3	-	3	-	-	3	-	-
Roche Elecsys 1010 / 2010	-	1	-	1	-	-	1	-	-
Roche Modular Analytics	-	2	-	2	-	-	2	-	-
Siemens ADVIA Centaur	-	2	-	1	1	-	2	-	-
Standard Diagnostics	-	1	-	-	1	-	1	-	-
VITROS ECI	-	1	-	1	-	-	1	-	-

<b><u>Method</u></b>	<b>Specimen VM-9</b>			<b>Specimen VM-10</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	55	2	-	55	2	-
Abbott Architect	27	-	-	27	-	-
bioMerieux Vidas, Mini Vidas	4	-	-	4	-	-
Roche cobas 6000 / e 601	13	-	-	13	-	-
Roche cobas e 411	3	-	-	3	-	-
Roche Elecsys 1010 / 2010	1	-	-	1	-	-
Roche Modular Analytics	2	-	-	2	-	-
Siemens ADVIA Centaur	2	-	-	2	-	-
Standard Diagnostics	-	1	-	-	1	-
VITROS ECI	1	-	-	1	-	-

**Viral Markers – HBeAg**

<b><u>Method</u></b>	<b>Specimen VM-6</b>			<b>Specimen VM-7</b>			<b>Specimen VM-8</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	-	39	-	-	39	-	-	39	-
Abbott Architect	-	16	-	-	16	-	-	16	-
bioMerieux Vidas, Mini Vidas	-	3	-	-	3	-	-	3	-
DiaSorin	-	1	-	-	1	-	-	1	-
Roche cobas 6000 / e 601	-	13	-	-	13	-	-	13	-
Roche cobas e 411	-	2	-	-	2	-	-	2	-
Roche Modular Analytics	-	2	-	-	2	-	-	2	-
Siemens ADVIA Centaur	-	1	-	-	1	-	-	1	-

<b><u>Method</u></b>	<b>Specimen VM-9</b>			<b>Specimen VM-10</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	1	38	-	-	39	-
Abbott Architect	-	16	-	-	16	-
bioMerieux Vidas, Mini Vidas	-	3	-	-	3	-
DiaSorin	-	1	-	-	1	-
Roche cobas 6000 / e 601	1	12	-	-	13	-
Roche cobas e 411	-	2	-	-	2	-
Roche Modular Analytics	-	2	-	-	2	-
Siemens ADVIA Centaur	-	1	-	-	1	-



**Viral Markers – Anti-HBs**

<b><u>Method</u></b>	<b>Specimen VM-6</b>			<b>Specimen VM-7</b>			<b>Specimen VM-8</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	117	2	-	13	103	3	5	113	1
Abbott Architect	50	-	-	6	43	1	1	49	-
Beckman ACCESS / 2 / Dxl bioMerieux Vidas, Mini Vidas	2	-	-	-	2	-	-	2	-
DiaSorin	1	-	-	-	1	-	-	1	-
Roche cobas 6000 / e 601	-	1	-	-	1	-	-	1	-
Roche cobas e 411	31	-	-	4	25	2	3	27	1
Roche Elecsys 1010 / 2010	11	-	-	1	10	-	1	10	-
Roche Modular Analytics	2	-	-	-	2	-	-	2	-
Siemens ADVIA Centaur	2	-	-	1	1	-	-	2	-
Standard Diagnostics	6	-	-	-	6	-	-	6	-
VITROS 3600/4600/5600	1	-	-	-	1	-	-	1	-
VITROS ECI	6	1	-	-	7	-	-	7	-
	3	-	-	1	2	-	-	3	-

<b><u>Method</u></b>	<b>Specimen VM-9</b>			<b>Specimen VM-10</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	116	3	-	115	3	-
Abbott Architect	50	-	-	50	-	-
Beckman ACCESS / 2 / Dxl bioMerieux Vidas, Mini Vidas	2	-	-	1	-	-
DiaSorin	1	-	-	1	-	-
Roche cobas 6000 / e 601	-	1	-	-	1	-
Roche cobas e 411	30	1	-	30	1	-
Roche Elecsys 1010 / 2010	11	-	-	11	-	-
Roche Modular Analytics	2	-	-	2	-	-
Siemens ADVIA Centaur	2	-	-	2	-	-
Standard Diagnostics	6	-	-	6	-	-
VITROS 3600/4600/5600	1	-	-	1	-	-
VITROS ECI	6	1	-	6	1	-
	3	-	-	2	-	-

**Viral Markers – HBsAg**

<b><u>Method</u></b>	<b>Specimen VM-6</b>			<b>Specimen VM-7</b>			<b>Specimen VM-8</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	1	170	-	5	166	-	4	167	-
Abbott Architect	-	70	-	-	70	-	-	70	-
Beckman ACCESS / 2 / Dxl bioMerieux Vidas, Mini Vidas	-	3	-	-	3	-	-	3	-
DiaSorin	1	2	-	1	2	-	1	2	-
Human	-	1	-	-	1	-	-	1	-
Roche cobas 6000 / e 601	-	34	-	1	33	-	1	33	-
Roche cobas e 411	-	18	-	1	17	-	1	17	-
Roche Elecsys 1010 / 2010	-	1	-	-	1	-	-	1	-
Roche Modular Analytics	-	3	-	1	2	-	1	2	-
Siemens ADVIA Centaur	-	7	-	-	7	-	-	7	-
Siemens Immulite/1000	-	1	-	-	1	-	-	1	-
Standard Diagnostics	-	8	-	-	8	-	-	8	-
VITROS 3600/4600/5600	-	7	-	-	7	-	-	7	-
VITROS ECI	-	4	-	1	3	-	-	4	-

<b><u>Method</u></b>	<b>Specimen VM-9</b>			<b>Specimen VM-10</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	4	167	-	1	169	-
Abbott Architect	-	70	-	-	70	-
Beckman ACCESS / 2 / Dxl bioMerieux Vidas, Mini Vidas	-	3	-	-	3	-
DiaSorin	1	2	-	1	2	-
Human	-	1	-	-	1	-
Roche cobas 6000 / e 601	1	33	-	-	34	-
Roche cobas e 411	1	17	-	-	18	-
Roche Elecsys 1010 / 2010	-	1	-	-	1	-
Roche Modular Analytics	1	2	-	-	3	-
Siemens ADVIA Centaur	-	7	-	-	7	-
Siemens Immulite/1000	-	1	-	-	1	-
Standard Diagnostics	-	8	-	-	8	-
VITROS 3600/4600/5600	-	7	-	-	7	-
VITROS ECI	-	4	-	-	3	-

**Viral Markers – Anti-HCV**

<b><u>Method</u></b>	<b>Specimen VM-6</b>			<b>Specimen VM-7</b>			<b>Specimen VM-8</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	-	150	-	150	-	-	28	118	4
Abbott Architect	-	69	-	69	-	-	1	68	-
Beckman ACCESS / 2 / Dxl	-	1	-	1	-	-	-	1	-
bioMerieux Vidas, Mini Vidas	-	3	-	3	-	-	-	3	-
Roche cobas 6000 / e 601	-	27	-	27	-	-	17	6	4
Roche cobas e 411	-	13	-	13	-	-	6	7	-
Roche Elecsys 1010 / 2010	-	1	-	1	-	-	1	-	-
Roche Modular Analytics	-	3	-	3	-	-	3	-	-
Siemens ADVIA Centaur	-	6	-	6	-	-	-	6	-
Standard Diagnostics	-	5	-	5	-	-	-	5	-
VITROS 3600/4600/5600	-	6	-	6	-	-	-	6	-
VITROS ECI	-	4	-	4	-	-	-	4	-

<b><u>Method</u></b>	<b>Specimen VM-9</b>			<b>Specimen VM-10</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	2	147	1	1	149	-
Abbott Architect	-	69	-	-	69	-
Beckman ACCESS / 2 / Dxl	-	1	-	-	1	-
bioMerieux Vidas, Mini Vidas	-	3	-	-	3	-
Roche cobas 6000 / e 601	-	26	1	-	27	-
Roche cobas e 411	1	12	-	-	13	-
Roche Elecsys 1010 / 2010	-	1	-	-	1	-
Roche Modular Analytics	-	3	-	-	3	-
Siemens ADVIA Centaur	1	5	-	1	5	-
Standard Diagnostics	-	5	-	-	5	-
VITROS 3600/4600/5600	-	6	-	-	6	-
VITROS ECI	-	4	-	-	4	-

**Toxoplasma gondii Antibody (IgG) - Qualitative**

<b><u>Method</u></b>	<b>Specimen TOX-3</b>			<b>Specimen TOX-4</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	1	18	-	19	-	-
Abbott Architect	1	10	-	11	-	-
bioMerieux Vidas, Mini Vidas	-	1	-	1	-	-
Roche cobas 6000 / e 601	-	1	-	1	-	-
Roche cobas e 411	-	2	-	2	-	-
Siemens ADVIA Centaur	-	1	-	1	-	-

**Toxoplasma gondii Antibody (IgG)—Quantitative (IU/mL)**

<b><u>Specimen/Method</u></b>	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>
<b>Specimen TOX-3</b>						
All Method	21	0.302	0.207	68.4	0.30	0.00 - 0.72
Abbott Architect	13	0.438	0.126	28.8	0.40	0.18 - 0.70
Roche cobas e 411	3	0.130	0.001	0.0	0.13	0.12 - 0.14
<b>Specimen TOX-4</b>						
All Method	20	453.870	267.086	58.8	350.95	0.00 - 988.05
Abbott Architect	14	391.157	244.294	62.5	200.00	0.00 - 879.75
Roche cobas e 411	3	801.000	230.128	28.7	906.00	340.74 - 1261.26

### Toxoplasma gondii Antibody (IgM) - Qualitative

<b><u>Method</u></b>	<b>Specimen TOX-3</b>			<b>Specimen TOX-4</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	-	22	-	21	-	-
Abbott Architect	-	13	-	12	-	-
bioMerieux Vidas, Mini Vidas	-	1	-	1	-	-
Roche cobas 6000 / e 601	-	2	-	2	-	-
Roche cobas e 411	-	3	-	3	-	-
Siemens ADVIA Centaur	-	1	-	1	-	-

### Toxoplasma gondii Antibody (IgM)—Quantitative (IU/mL)

<b><u>Specimen/Method</u></b>	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>
<b>Specimen TOX-3</b>						
ALL METHODS	19	0.161	0.039	24.0	0.17	0.08 - 0.24
Abbott Architect	12	0.148	0.035	23.7	0.16	0.07 - 0.22
<b>Specimen TOX-4</b>						
ALL METHODS	18	10.160	6.787	66.8	7.67	0.00 - 23.74
Abbott Architect	12	7.501	0.355	4.7	7.53	6.79 - 8.22

**Cytomegalovirus (CMV) Antibodies (IgG) - Qualitative**

<b><u>Method</u></b>	<b>Specimen CMV-3</b>			<b>Specimen CMV-4</b>		
	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>	<b><u>Positive</u></b>	<b><u>Negative</u></b>	<b><u>Equivocal</u></b>
ALL METHODS	2	11	-	-	13	-
Abbott Architect	2	9	-	-	11	-
Roche cobas 6000 / e 601	-	2	-	-	2	-

**Cytomegalovirus (CMV) Antibodies (IgG) —Quantitative (U/mL)**

<b><u>Specimen/Method</u></b>	<b><u>Labs</u></b>	<b><u>Mean</u></b>	<b><u>SD</u></b>	<b><u>CV</u></b>	<b><u>Median</u></b>	<b><u>Range</u></b>
<b>Specimen CMV-3</b>						
All Method	16	3.521	1.892	53.7	4.30	0.00 - 7.31
Abbott Architect	12	4.542	0.582	12.8	4.70	3.37 - 5.71
<b>Specimen CMV-4</b>						
All Method	16	1.247	1.102	88.4	0.85	0.00 - 3.46
Abbott Architect	12	1.600	1.054	65.9	1.35	0.00 - 3.71

### Cytomegalovirus (CMV) Antibodies (IgM) - Qualitative

<u>Method</u>	Specimen CMV-3			Specimen CMV-4		
	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>	<u>Positive</u>	<u>Negative</u>	<u>Equivocal</u>
ALL METHODS	18	-	-	-	18	-
Abbott Architect	13	-	-	-	13	-
DiaSorin	1	-	-	-	1	-
Roche cobas 6000 / e 601	3	-	-	-	3	-
Roche cobas e 411	1	-	-	-	1	-

### Cytomegalovirus (CMV) Antibodies (IgM)—Quantitative (U/mL)

<u>Specimen/Method</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
<b>Specimen CMV-3</b>						
All Method	13	3.065	2.576	84.0	2.09	0.00 - 8.22
Abbott Architect	11	2.016	0.300	14.9	2.07	1.41 - 2.62
<b>Specimen CMV-4</b>						
All Method	13	0.238	0.095	40.0	0.19	0.04 - 0.43
Abbott Architect	11	0.247	0.102	41.1	0.23	0.04 - 0.46

### Glycohemoglobin (percent)

<u>Method</u>	Specimen GH-3						Specimen GH-4					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	194	7.74	0.34	4.4	7.8	7.2 - 8.3	204	5.71	0.26	4.5	5.7	5.3 - 6.1
All Hemoglobin A1c Methods	194	7.74	0.34	4.4	7.8	7.2 - 8.3	204	5.71	0.26	4.5	5.7	5.3 - 6.1
All TOSOH Methods	19	7.78	0.19	2.4	7.8	7.3 - 8.3	19	5.61	0.10	1.7	5.6	5.2 - 6.0
Beckman AU A1c	13	7.86	0.42	5.3	7.8	7.3 - 8.4	13	5.63	0.28	5.0	5.6	5.2 - 6.0
Bio-Rad D-100	11	10.98	1.31	11.9	10.7	10.3 - 11.7	17	6.32	0.19	3.1	6.4	5.9 - 6.7
Roche cobas c 501 HbA1c	11	7.29	0.27	3.7	7.3	6.8 - 7.8	11	5.34	0.17	3.3	5.4	5.0 - 5.7
Siemens DCA Vantage	82	7.72	0.24	3.1	7.8	7.2 - 8.2	82	5.80	0.17	2.9	5.8	5.4 - 6.2
Siemens Dimension HB1C	28	7.73	0.27	3.5	7.8	7.2 - 8.2	29	5.58	0.13	2.4	5.6	5.2 - 6.0
TOSOH G8	19	7.78	0.19	2.4	7.8	7.3 - 8.3	19	5.61	0.10	1.7	5.6	5.2 - 6.0

**CK-MB - Quantitative (U/L)**

<u>Specimen/Method</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
<b>Specimen CK-6</b>						
All Method	1	-	-	-	83.1	Not graded
<b>Specimen CK-7</b>						
All Method	1	-	-	-	15.8	Not graded
<b>Specimen CK-8</b>						
All Method	1	-	-	-	25.8	Not graded
<b>Specimen CK-9</b>						
All Method	1	-	-	-	45.2	Not graded
<b>Specimen CK-10</b>						
All Method	1	-	-	-	7.6	Not graded

**Medical Laboratory Evaluation**  
25 Massachusetts Ave NW Ste 700  
Washington, DC 20001-7401  
800-338-2746 • 202-261-4500 • Fax: 202-835-0440  
*www.acponline.org/mle*